

Marquette University

e-Publications@Marquette

Biomedical Engineering Faculty Research and
Publications

Biomedical Engineering, Department of

3-2012

The Latest Lessons Learned from Retrieval Analyses of Ultra-High Molecular Weight Polyethylene, Metal-on-Metal, and Alternative Bearing Total Disc Replacements

Steven M. Kurtz
Drexel University

Jeffrey M. Toth
Marquette University, jeffrey.toth@marquette.edu

Ryan Siskey
Drexel University

Lauren Ciccarelli

Daniel MacDonald
Drexel University

See next page for additional authors

Follow this and additional works at: https://epublications.marquette.edu/bioengin_fac



Part of the [Biomedical Engineering and Bioengineering Commons](#)

Recommended Citation

Kurtz, Steven M.; Toth, Jeffrey M.; Siskey, Ryan; Ciccarelli, Lauren; MacDonald, Daniel; Isaza, Jorge; Lanman, Todd; Punt, Ilona; Marla Steinbeck, Marla; Goffin, Jan; and Ooij, André van, "The Latest Lessons Learned from Retrieval Analyses of Ultra-High Molecular Weight Polyethylene, Metal-on-Metal, and Alternative Bearing Total Disc Replacements" (2012). *Biomedical Engineering Faculty Research and Publications*. 67.

https://epublications.marquette.edu/bioengin_fac/67

Authors

Steven M. Kurtz, Jeffrey M. Toth, Ryan Siskey, Lauren Ciccarelli, Daniel MacDonald, Jorge Isaza, Todd Lanman, Ilona Punt, Marla Marla Steinbeck, Jan Goffin, and André van Ooij

Marquette University

e-Publications@Marquette

Biomedical Engineering Faculty Research and Publications/College of Engineering

This paper is NOT THE PUBLISHED VERSION; but the author's final, peer-reviewed manuscript. The published version may be accessed by following the link in the citation below.

Seminars in Spine Surgery, Vol. 24, No. 1 (March 2012): 57-70. [DOI](#). This article is © Elsevier and permission has been granted for this version to appear in [e-Publications@Marquette](#). Elsevier does not grant permission for this article to be further copied/distributed or hosted elsewhere without the express permission from Elsevier.

The Latest Lessons Learned from Retrieval Analyses of Ultra-High Molecular Weight Polyethylene, Metal-on-Metal, and Alternative Bearing Total Disc Replacements

Steven M. Kurtz

Exponent Inc., Philadelphia, PA

School of Biomedical Engineering, Science, and Health Systems Drexel University, Philadelphia, PA

Jeffrey M. Toth

Department of Orthopaedic Surgery, Medical College of Wisconsin, Milwaukee, WI

Ryan Siskey

Exponent Inc., Philadelphia, PA

School of Biomedical Engineering, Science, and Health Systems Drexel University, Philadelphia, PA

Lauren Ciccarelli

Exponent Inc., Philadelphia, PA

Daniel MacDonald

Exponent Inc., Philadelphia, PA

School of Biomedical Engineering, Science, and Health Systems Drexel University, Philadelphia, PA

Jorge Isaza

Our Lady of the Lake Regional Medical Center, Baton Rouge, LA

Todd Lanman

Olympia Medical Center, Beverly Hills, CA

Ilona Punt

Department of Orthopaedic Surgery, Maastricht University Medical Center, Maastricht, The Netherlands

Research School CAPHRI, Maastricht University, Maastricht, The Netherlands

Marla Steinbeck

School of Biomedical Engineering, Science, and Health Systems Drexel University, Philadelphia, PA

Jan Goffin

Department of Neurosurgery, University Hospitals Leuven, KU Leuven, Leuven, Belgium

Andrévan Ooij

Department of Orthopaedic Surgery, Viecuri Medical Center, Venlo, The Netherlands

Knowledge regarding the in vivo performance and periprosthetic tissue response of cervical and lumbar total disc replacements (TDRs) continues to expand. This review addresses the following 4 main questions: (1) What are the latest lessons learned from using polyethylene in large joints and how are they relevant to current TDRs? (2) What are the latest lessons learned regarding adverse local tissue reactions from metal-on-metal cobalt-chrome bearings in large joints and how are they relevant to current TDRs? (3) What advancements have been made in understanding the in vivo performance of alternative biomaterials, such as stainless steel and polycarbonate urethane, for TDRs in the past 5 years? (4) How has retrieval analysis of all these various artificial disc bearing technologies advanced the state-of-the-art in preclinical testing of TDRs? The study of explanted artificial discs and their associated tissues can help inform bearing selection as well as the design of future generations of disc arthroplasty. Analyzing retrieved artificial discs is also essential for validating preclinical test methods.

Keywords

adverse local tissue reactions, metal-on-metal, total disc replacement, retrieval analysis, UHMWPE, wear

Knowledge regarding the in vivo performance and periprosthetic tissue response of cervical and lumbar total disc replacements (TDRs) continues to expand. Many different cervical and lumbar artificial discs are currently available and are fabricated from a range of polymers and metals, which may be unfamiliar to a spine surgeon or resident. Some of these biomaterials, such as ultra-high molecular weight polyethylene (hereafter, polyethylene) and cobalt-chromium alloy (hereafter, CoCr alloy), have a long-standing history as orthopedic bearing materials for total joint replacements.¹ Artificial disc designers have also incorporated novel biomaterials as bearing materials for artificial discs,² including titanium alloys, stainless steel, polycarbonate urethanes (PCUs), and polyaryletheretherketone. Until recently, it has remained unclear which bearing materials are sufficiently durable, biocompatible, and resistant to mechanical loading for use in cervical or lumbar disc arthroplasty. As noted in previous reviews,^{3, 4} the study of explanted artificial discs and periprosthetic tissues can help inform bearing selection decisions as well as improve the design of future generations of disc arthroplasty.

Bearing selection continues to be an important topic for large total joint replacements used in the hip and the knee.⁵ In response to concerns regarding wear, osteolysis, and instability of polyethylene acetabular liners, major evolutionary changes to the biomaterials and bearing designs used for hip arthroplasty have occurred over the past decade.¹ In the late 1990s, the vast majority of hip replacements performed in the United States included gamma-sterilized polyethylene liners, whereas today 4 types of bearing materials are in clinical use (Table 1). Highly crosslinked polyethylene is the most widely used hip bearing material for the acetabulum; however, alternative bearings such as metal-on-metal (MOM) and ceramic-on-ceramic were, for a while, increasingly adopted in the mid-2000s.⁶ These premium “hard-on-hard” bearings are more expensive and sensitive to surgical positioning than highly crosslinked polyethylene,^{6, 7} and, consequently, their use has since waned. Of the contemporary bearing material combinations used for large total joint replacement, only metal-on-polyethylene and MOM articulations have been used for disc replacements. One of the goals of this review is to synthesize the latest developments in bearing technology from orthopedics with the emerging perspective obtained from analyzed retrieved TDRs (Table 1).

Table 1. Total Hip Replacement Bearing Materials¹

Type	Femoral Head	Acetabular Liner
M-PE	CoCr alloy	Polyethylene
C-PE	Ceramic	Polyethylene
COC	Ceramic	Ceramic
MOM	CoCr alloy	CoCr alloy

In addition to informing bearing selection, analyzing retrieved artificial discs is also essential for validating preclinical test methods.⁸ As we gain experience with the clinical performance of artificial discs, the observations of retrieved implants provide valuable insight for the methodologies used in wear and fatigue testing.⁸ Working with regulatory agencies under the auspices of the American Society for Testing and Materials International (ASTM), biomedical engineers and surgeons have helped to advance the state-of-the-art in wear testing, not only under normal intended use conditions, but also under more demanding, impingement scenarios that are frequently observed in retrieved artificial discs.

None of the advancements described in this review would have been possible without the close and active collaboration between surgeons and biomedical engineers. In recent years, researchers have been studying explanted artificial discs and periprosthetic tissues from retrieved cervical and lumbar artificial discs to better understand the durability and biocompatibility of these implants.^{3, 4} Many of these artificial discs and tissues were collected as part of a public, multicenter, federally funded retrieval research program,^{3, 4} but retrieval studies are also conducted by manufacturers as part of Food and Drug Administration (FDA)-mandated postmarket surveillance programs. This article builds on previous reviews that focused on the motivation, methodologies, and early findings from the first retrieval analyses of artificial discs.^{3, 4} In this article, we concentrate on addressing the following 4 main questions: (1) What are the latest lessons learned from using polyethylene in large joints and how are they relevant to current TDRs? (2) What are the latest lessons learned regarding adverse local tissue reactions from MOM CoCr bearings in large joints and how are they relevant to current TDRs? (3) What advancements have been made in understanding the in vivo performance of alternative biomaterials, such as stainless steel and PCU, for TDRs in the past 5 years? (4) How has retrieval analysis of all these various artificial disc bearing technologies advanced the state-of-the-art in preclinical testing of TDRs?

Update on Polyethylene Used in Total Joint Replacements and Disc Replacement

In total hip and knee arthroplasty, the dominant bearing technology for the past 50 years continues to be a CoCr alloy femoral component articulating against a polyethylene acetabular or tibial component.¹ Especially within the past 2 decades, the polyethylene used in hip and knee joint arthroplasty has undergone major changes in terms of formulation.⁵ By contrast, the changes to polyethylene formulation in the spine are comparatively subtle, mirroring the early evolution of sterilization practices for polymer hip and knee components in the early 1990s.⁹ Before the mid-1990s, polyethylene was gamma sterilized in air-permeable packaging, which contributed to shelf aging before implantation and in vivo oxidation. In the 1990s, the state-of-the-art for polyethylene shifted to gamma sterilization in a low-oxygen environment (often referred to as “gamma inert sterilization”). Gamma sterilization of polyethylene in air was, therefore, gradually abandoned, but variation in industry practices for packaging limited the effectiveness of this transition.¹⁰ Indeed, the first generation of gamma sterilization in nitrogen by Link, the original producers of the CHARITÉ TDR, were permeable to oxygen and, therefore, exhibited similar degradation as earlier implants that were gamma sterilized in air.⁹ Oxygen-impermeable packaging is the key to successful preservation of polyethylene properties during shelf aging.¹⁰

Beginning in 1998, several generations of highly crosslinked polyethylene materials have been introduced and have since become the standard of care in hip arthroplasty.¹ It is now clear from the orthopedic literature that highly crosslinked polyethylene reduces wear as well as the risk of osteolysis in total hip replacements (THRs).⁵ However, not all implant designs may be appropriate for highly crosslinked polyethylene. Although elevated crosslinking improves wear resistance, it also reduces the material's ductility and fracture resistance.¹¹ The concomitant decrease in mechanical properties that occurs during the crosslinking process may be deleterious to both knee and spine devices. Thus, gamma-inert-sterilized polyethylene continues to be used in total knee and disc replacements, at least partly because of the increased stresses that may be encountered in these applications.

Much of the existing retrieval evidence published in journal articles thus far for polyethylene in TDRs is based on the equivalent of historical gamma-air-sterilized material that was abandoned almost 2 decades ago. Why, then, continue to study these early polyethylene TDRs? It turns out that these explants helped to dispel many of the optimistic preconceptions offered by the early promoters of disc arthroplasty. In 2003, Link and Keller¹² wrote that “it is reassuring that the anterior column of the lumbar spine appears to be one place in the human body where periprosthetic osteolysis is not a major factor, due to lower ranges of motion and an absence of synovium compared to the hip and knee joint.” During the early 2000s, the issues of wear and osteolysis were poorly understood and, hence, underestimated by the spine community advocating disc arthroplasty as an alternative to fusion.

The first systematic retrieval analyses of a collection of polyethylene TDR components, performed on the CHARITÉ artificial disc, was instrumental in challenging the established paradigm in the spine community that wear was irrelevant for disc arthroplasty. A series of disc retrieval and clinical publications helped to establish osteolysis and polyethylene wear as clinically relevant, although rare, complications of disc arthroplasty.^{13, 14, 15} On examination, the central dome and peripheral rim of the explanted polyethylene components exhibited different wear and damage mechanisms.¹⁶ The central dome often exhibited burnishing and the microscopic, multidirectional scratching consistent with the adhesive and abrasive wear mechanism characteristic of THR. On the other hand, the peripheral rim of the retrievals sometimes showed evidence of delamination, cracking, and fracture, damage modes most often associated with gamma-air-sterilized total knee replacements. The central dome adhesive/abrasive wear was associated with the expected, dome-to-dome articulation of the disc arthroplasty, whereas the peripheral rim cracking and fractures were associated with

impingement. These basic observations, performed on a collection of 21 retrieved CHARITÉ TDRs,¹⁶ would provide the basis for starting to validate the methods used to test artificial discs.⁸ Subsequent publications focused on quantitatively characterizing the patterns of wear¹⁷ and the natural history of oxidative degradation⁹ of the polyethylene TDRs.

Having demonstrated that clinically relevant wear was rare, but indeed possible, with historical gamma-air-sterilized polyethylene TDRs, researchers turned to study the characteristics and biological ramifications of wear debris released into the intradiscal periprosthetic tissues.^{18, 19, 20} In one of the only long-term studies of retrieved TDR tissue, we observed a chronic inflammatory response within the periprosthetic fibrous tissues from 15 of 16 patients who had undergone revision surgery of historical CHARITÉ TDRs.²⁰ Revision for all these patients was indicated for intractable pain after an average of 9 years implantation (range: 3-16 years). The inflammatory cells consisted of lymphocytes, macrophages, and giant cells, which were associated with ingested small or uningested large polyethylene particles. Polyethylene particles $>2\ \mu\text{m}$ were detected in tissue samples from 15 of 16 patients. In addition, the presence of giant cells and polyethylene wear debris increased with implantation time, and was associated with inflammatory cytokines such as tumor necrosis factor- α and interleukin-6 in a subset of CD-68-positive macrophages and giant cells. The highest number of polyethylene particles was observed in the single patient who showed signs of osteolysis of the sacrum.^{14, 19, 20} Innervation and vascularization were also noted within the retrieved tissue. Inflammation and innervation may contribute to the development of neuroinflammatory-induced pain in TDR patients. These findings point to the complexity of the wear debris' interactions in the spine and of the clinically relevant wear debris required to stimulate histiocytes and giant cell formation.

In total hip arthroplasty tissues, submicron-sized polyethylene particles have been shown to be critical in promoting the phagocytic inflammatory response and stimulating the production of proinflammatory factors. Further evaluation of 5 TDR patient tissues with wear and inflammation by scanning electron microscopy (SEM) showed more than a billion 0.5- to 2.0- μm polyethylene wear particles per gram of tissue ($1.6 \times 10^9/\text{g}$).¹⁹ Compared with gamma-air-sterilized total hip arthroplasty revision tissue ($2.3 \times 10^9/\text{g}$), a significantly lower concentration of particles was found in TDR tissue; however, this represents a substantial load within spinal tissue. Moreover, despite differences in loading and kinematics between the lumbar spine and the hip joint, the mean wear particle size and shape were comparable, although the TDR particles tended to be smaller and more round. No correlations were found between visible damage to the polyethylene core and the concentration or shape of the polyethylene particles. A positive correlation was found for increasing particle size and increasing rim penetration of the TDR core. The lower concentration of polyethylene particles in TDR tissue might explain why pain, rather than osteolysis, is the major reason for revision surgery. For the single osteolytic TDR patient and sixth patient evaluated, a high concentration of particles was detected ($2.58 \times 10^9/\text{g}$).

Based on the data from retrieved historical TDRs and their periprosthetic tissues, it appears that, overall, the clinical consequences of gamma-air-sterilized polyethylene have conformed to expectations from hip and knee arthroplasty, with a few twists or “surprises.” Perhaps the first surprise is that osteolysis is so rare after these historical TDRs, especially because it is clear the periprosthetic particle load and characteristics are at least comparable with hip arthroplasty in some patients. We observed on the order of 1 billion polyethylene particles per gram of the retrieved periprosthetic tissues from these TDR patients, yet—with only 1 exception in our series—these levels were not associated with osteolysis of the adjacent vertebral bodies.¹⁹ The second unexpected finding relates to observations of impingement, which were common in the collection of retrieved disc replacements. In addition, we observed new nerve growth in certain samples of the periprosthetic tissues. It remains unclear at present whether there is an association between wear particles and the discogenic pain that was the dominant reason for revision of disc arthroplasties incorporating historical polyethylene.

New data are beginning to emerge regarding conventional gamma-inert-sterilized polyethylene in cervical and lumbar TDRs. The largest collection of ProDisc retrievals, for example, has been maintained by researchers from the Hospital for Special Surgery (HSS), who have presented their observations over the past 5 years in a series of conference abstracts.^{21, 22, 23, 24} In the most recent analysis from HSS,²¹ 20 lumbar ProDisc retrievals, implanted for an average of 1 year, showed evidence of impingement in 15 of 20 (75%) and backside wear in 11 of 14 (78%) of the components that were disassembled. In a companion HSS study²³ of 29 cervical ProDisc retrievals, also implanted for an average of 1 year, researchers observed impingement in 96% (28 of 29) of explanted devices. These short-term retrieval findings, coupled with radiographic evidence of impingement in a clinical study²⁵ of the lumbar ProDisc, further highlight the relevance of impingement to contemporary TDR designs in addition to the historical CHARITÉ. In our retrieval collection, we have had the opportunity to characterize 10 TDRs from conventional gamma-inert-sterilized polyethylene. These components, implanted for 0.3-3.3 years, exhibited an order of magnitude lower oxidation than the historical gamma-air-sterilized materials analyzed previously. Two of these devices could not be fully examined owing to severe iatrogenic damage; however, we observed evidence of impingement on either the polyethylene components or the metallic endplates in all the remaining devices. Longer-term retrievals are necessary to effectively compare wear rates and oxidative properties between TDR designs fabricated from historical gamma-air-sterilized and conventional gamma-inert-sterilized polyethylene.

From a tissue response perspective, compared with the response to gamma-air-sterilized polyethylene, less has been published regarding the response to wear debris or inflammation in periprosthetic spine tissue from conventional gamma-inert-sterilized polyethylene TDRs. From our own studies, we have observed early tissue responses after a lumbosacral disc replacement with a ProDisc-L implant, which was removed after 14 months for low-back pain.²⁶ Microcomputed tomography of the tissue showed the presence of third-body debris, which was predominantly heterotopic ossification/bone formation. Histological analysis showed fibrotic tissue with increased vascularization. Other areas of the tissue showed evidence of cell degeneration, and several fields contained fibrocartilage. Similar to previous short-term studies, no phagocytic cells were observed in areas where polyethylene wear debris was found and the amount of wear debris was limited.

Taken together, although limited, the available data provide support for gamma-inert-sterilized polyethylene in disc arthroplasty. Research on historical gamma-air-sterilized polyethylene TDRs continues, as it provides an ideal negative control material against which modern polyethylene materials for the spine can be compared and effectively benchmarked. As we shall see in the following sections, the foundation of knowledge gained from polyethylene TDR retrievals forms a strong basis for critically examining, and indeed expanding, the battery of preclinical tests new TDRs must be subjected to before their widespread clinical adoption.

Adverse Local Tissue Reactions in CoCr Alloy MOM TDR

Until recently, CoCr alloy MOM hip bearings were considered to be a reasonable alternative to reduce long-term wear and short-term dislocation risk in hip arthroplasty, even though long-term exposure to metal ions has been a concern with MOM for many decades.²⁷ However, there has been a recent increase in the reports in the hip literature of short-term revision due to metal hypersensitivity, osteolysis, and pseudotumor formation.^{28, 29} These adverse local tissue reactions to MOM, although infrequent, may result in substantial tissue damage and permanent disability to hip patients. In July 2008, Zimmer issued a voluntary recall for the Durom MOM hip system (Zimmer, Warsaw, IN) because its “warnings and instructions for use were inadequate.”³⁰ In April 2010, the British Medicines and Healthcare products Regulatory Agency (MHRA) (equivalent to the FDA) issued a Medical Device Alert for all MOM THRs and called for systems to be put into place to monitor patients implanted with MOM for soft-tissue reactions and unexplained pain.³¹ In August 2010, DePuy Orthopaedics issued a voluntary worldwide recall for ASR MOM hip implant systems, owing to a

higher-than-expected revision rate in the United Kingdom.³⁰ In May 2011, the FDA ordered orthopedic manufacturers to perform postmarket surveillance studies of MOM hip implants in the United States and established a Web site³⁰ to inform the public and encourage the reporting of adverse events related to MOM bearings. Media articles questioning the appropriateness of MOM hip bearings and mass tort litigation have since followed.^{32, 33} Consequently, the short-term revision due to periprosthetic tissue reactions of MOM, heightened scrutiny by the lay press, the prospect of involvement in litigation, and the recent regulatory actions have reduced the level of enthusiasm for this alternative hip bearing in the United States. Spine surgeons considering to use MOM TDRs should familiarize themselves with the FDA Web site for MOM hip implants,³⁰ which will be updated regularly as new information becomes available.

The MOM situation for TDR is more complex than for THR. A unique problem for spinal surgeons in understanding the tissue response to particulate debris, metal ions, and corrosion products associated with “Metal-on-Metal” TDR is that—unlike in hip arthroplasty—a variety of alloys are used in artificial discs.² For example, CoCr alloy is used in the Maverick Lumbar disc (Medtronic Spinal and Biologics, Memphis, TN), the Flexicore cervical and lumbar discs (Stryker Spine, Allendale, NJ), and in the Kineflex cervical and lumbar discs (Spinal Motion, Mountain View, CA). Stainless steel is used in the Prestige cervical disc. Titanium alloy is used in the Prestige LP cervical disc. Finally, commercially pure titanium is used in the titanium shells of the Bryan artificial disc (although the titanium shells do not articulate in normal circumstances, the shells may produce wear debris if impingement or failure occurs). Thus, the particulate debris, metal ions, and corrosion products, as well as the attendant host response found in periprosthetic tissues, varies with the specific alloy used in the device. Therefore, the tissue response seen for a device made of 1 alloy may not be the same as a tissue response for another alloy.

We have studied 7 explanted Maverick TDRs (Medtronic, Memphis, TN). The explanted Maverick devices were made available and funded through Medtronic's IDE studies or Medtronic's quality system. The Maverick is machined from cobalt-chromium-molybdenum alloy hot-worked bar stock (ASTM F-1537, Alloy 2). Protruding from the bone-abutting surface is a “keel” that provides for device stability by press-fitting into a prepared channel in the vertebral body. The articulating surface has a male (convex) dome that mates with the female (concave) feature of the superior endplate.

The 7 retrieved Maverick discs in our collection were removed from L4/L5 to L5/S1 after 0.6-3.1 years of implantation (average: 1.3 years). All the retrieved components appeared polished but showed varying extent of scratches on the bearing surfaces and rims (Fig. 1). There was no macroscopic evidence of plastic deformation, third-body wear, burnishing, pitting, or fracture at either the rims or the spherical bearing surfaces of the implants. The primary wear mechanism was microabrasion, which was evident by microscopic scratching of the articulating surfaces in all 3 sets of components. Two sets of components, implanted for 0.6 and 1.5 years, showed evidence of rim impingement. We found a fan-shaped pattern of microscopic scratches on both the superior and inferior endplates as clear evidence of chronic impingement in this case (Fig. 1). Located in the anterior aspect of the device, the impingement zones consisted of microscopic, unidirectional, circumferential scratches, suggesting that they were generated by a combination of lateral bending and/or axial rotation. We also consistently found surface deposits, manifested as a smoky or hazy discoloration, on the superior and inferior endplates of all the retrievals. Using low-voltage EDS analysis, we confirmed that these carbon- and oxygen-rich films were of comparable composition with those tribochemical reaction layers previously observed in well-functioning MOM hip joints.³⁴ We are continually collecting retrieved TDRs to understand long-term outcomes and potential complications.




Figure 1. Micrograph of the superior (left) and inferior (right) endplates of a Maverick retrieved implant (1.5 years in vivo) showing scratches and an area of impingement. (Color version of figure is available online.)

The periprosthetic tissues associated with one of the retrieved Maverick discs in our collection were also available for analysis. Periprosthetic tissues associated with retrieved Maverick devices have shown focal metallosis on gross examination. A uniform discoloration (metallosis) was not observed in any of the tissue samples at gross examination. Two minute metallic fragments were observed (as detected by Faxitron high-resolution radiography) remaining in the tissue samples after removal of bone subsequent to decalcification. As seen in Figure 2, Figure 3, histology showed that necrotic bone, necrotic bone marrow, and necrotic dense fibrous connective tissues were seen in all 3 periprosthetic tissue samples. As seen in Fig. 3, focal microscopic metallic debris was only infrequently observed in the histology of some of the tissue samples. Metallic debris was not found in a uniform distribution throughout the tissue samples. Unlike histology from the Prestige devices we previously reported on,^{4, 35} corrosion products were not observed in any of the tissue samples from the current Maverick TDR patient. Cytological (morphologic) changes associated with tissue necrosis were observed in the histology. As seen in Fig. 4, karyopyknosis, the irreversible condensation of chromatin in the nucleus of a cell undergoing necrosis or apoptosis, was observed in the histology. Karyorrhexis, or fragmentation of the pyknotic nucleus with subsequent accumulation of nuclear dust (heterochromatin nuclei of dead cells), was also observed in the histology. Karyolysis, the complete dissolution of the nucleus, was also observed in the histology. The cause of the tissue necrosis was not known. There was no evidence in the histology to suggest that an infection caused the necrosis. However, a focal cell-mediated immune response consisting of 25+ lymphocytes and <10 macrophages per 400× to 500× high-powered field was focally observed in periprosthetic tissues. The observed host response did not have an acute inflammatory character (no neutrophils). There were no signs of infection in the tissues. Osteolysis was not observed in any of the tissues. Based on the population of lymphocytes observed, there is concern that cobalt TDRs, such as the Maverick device, may be associated with tissue necrosis and lymphocytic (cell-mediated immune) response in some patients.




Figure 2. Histology of periprosthetic tissues adjacent to a cobalt-based TDR showing necrotic bone with empty osteocyte lacunae and necrotic marrow (H&E, original magnification = 70×). (Color version of figure is available online.)




Figure 3. Histology of periprosthetic tissues adjacent to a cobalt-based TDR showing necrotic tissue and microscopic metallic debris (H&E, original magnification = 79×). (Color version of figure is available online.)




Figure 4. Histology of periprosthetic tissues adjacent to a cobalt-based TDR showing karyopyknosis, karyorrhexis, and nuclear dust in tissues undergoing necrosis (H&E, original magnification = 500×). (Color version of figure is available online.)

A paucity of information on the host response to CoCr MOM TDRs exists in the literature. In fact, there are only 3 articles in the recent literature that present histologic findings associated with CoCr MOM TDRs.^{36, 37, 38} In addition to the tissue response we described earlier, periprosthetic tissue reactions to both fixed and mobile bearing CoCr MOM TDRs are known in both the cervical and lumbar spine. The article by Berry et al³⁸ indicated that “the final pathology report described the tissue as benign and reactive and consistent with a large granuloma. There were uniformly sized, shaped, and colored particles within the granulation tissue, indicative of wear debris.” However, Fig. 3 in the article by Berry et al³⁸ shows similar findings to the periprosthetic tissues we described earlier—profound tissue necrosis with either nuclear dust or metallic debris present. Similarly, a recent article by Guyer et al reported on a single Maverick lumbar disc prosthesis case report.³⁶ The authors state, “A laparoscopic biopsy found slightly necrotic nonvascular grayish-white tissue. Histology showed predominantly necrotic fibrous and adipose tissue. Bordering the necrosis were some focal, poorly defined, histiocytic palisades, and scattered foreign-body giant cells with surrounding lymphocytic infiltrate.”³⁶ Based on the tissue response we described earlier, as well as the 2 single case reports in the literature, there is concern that cobalt TDRs, such as the Maverick device, may be associated with tissue necrosis and lymphocytic (cell-mediated immune) response in some patients, similar to THRs. Neither the extent of this response nor the cause of this response can be gauged based on these few case reports. In summary, there is evidence of adverse local tissue reactions to CoCr alloy MOM TDRs, but these are thus far limited to suspected³⁹ or histologically confirmed^{36, 37, 38} cases in isolated reports in both cervical and lumbar disc applications.

Update on Device and Tissue Response to Alternative Biomaterials Used in TDR

Although polyethylene and CoCr alloy have a well-established track record as orthopedic bearing materials that can be traced back over 5 decades, as noted in the previous section, many different polymers and metal alloys have been incorporated into modern TDRs, especially in the cervical spine. These “alternative biomaterials” have no long-term clinical history as bearing materials, and hence, retrieval analysis has the opportunity to provide especially crucial feedback regarding the tribological performance, chemical stability, and tissue response to these new materials in the context of spine arthroplasty. In this section, we provide an update on our device and tissue retrieval experience for some of these new biomaterials used in TDRs. Because these alternative biomaterials are isolated to unique TDR designs, the following sections are grouped by device.

Stainless Steel—Prestige Cervical Disc System

The use of stainless steel as a bearing material in cervical arthroplasty can be traced to the Bristol/Cummins Disc, which was first used clinically in 1991. Early clinical prototypes of stainless steel cervical discs were evaluated in the late 1990s, leading to development of the Prestige ST (currently referred to as the Prestige Cervical Disc System) design by Medtronic Spinal and Biologics in 2002. This implant design has 8 components—2 endplates that interface via a ball-and-trough mechanism, 2 bone screws per endplate (4 bone screws in total), and 1 set screw per endplate (2 set screws in total). A grit-blasted surface on the endplates of the superior and inferior components allows for bone ongrowth. The spherical superior endplate is composed of a convex (ball) geometry that articulates with concave (trough) geometry of the inferior endplate. When the device is well fixed and functioning normally, articulation occurs between the ball of the superior endplate and the trough of the inferior endplate. However, if one or both of the endplates become loose or subside, then articulation can occur not only between the rims of the endplates, but also between the endplates and screws. In the United States, the FDA approved the Prestige cervical artificial disc in 2007. This TDR is intended to replace a cervical disc from C3 to C7 after removal of the disc for intractable radiculopathy and/or myelopathy, and it is implanted using an anterior surgical approach.

We have studied 20 explanted Prestige TDRs from 20 patients after an average implantation time of 2.0 years (0.3-7.0 years). The explanted devices were made available and funded through Medtronic's IDE studies or Medtronic's quality system. The explants generally exhibit a slightly discolored, elliptical wear region of varying dimension centered in the bearing center, with the long-axis oriented in the medial-lateral direction (Fig. 5). Microabrasive wear is the dominant in vivo wear mechanism, with microscopic scratches generally oriented in the medial-lateral direction.



Figure 5. Micrograph of the superior (left) and inferior (right) endplates of a Prestige retrieved implant (2.3 years in vivo) showing a typical wear scar.

Evidence of anterior impingement has been noted in the Prestige retrieval collection (11/16 explants, 69%) (Fig. 6). Evidence of localized screw hole fretting and fretting near the heads of bone screws is also typically observed, but SEM examination has ruled out corrosive material removal at these interfaces. We have seen evidence of locking screw fracture (3 cases) and bone screw fracture (1 case) among our retrievals. We continue to collect Prestige retrievals to better understand long-term outcomes and potential complications.




Figure 6. Anterior impingement in a Prestige retrieved implant (2.2 years in vivo). (A) Photomicrograph of entire component; (B) detail of impingement region.

From a periprosthetic tissue perspective, we have analyzed samples from approximately 15 Prestige cervical artificial discs that were explanted between 1 and 7 years. At gross examination, periprosthetic tissue samples from Prestige explants typically show focal metallosis. At longer postimplantation periods, a uniform discoloration (metallosis) has been observed in periprosthetic tissue samples at gross examination. High-resolution radiographs of periprosthetic tissues show the presence of infrequent small metallic debris before and after decalcification. Fibrous connective tissues are frequently found in the histology associated with Prestige explants. Focal microscopic metallic debris is observed histologically in periprosthetic tissue samples. Metallic debris is not found in a uniform distribution throughout the tissue samples. As exemplified in Fig. 7, in most periprosthetic tissues associated with Prestige explants, a lamellar distribution of metallic debris within the interstices of fibrous tissues is observed, especially at the device interface. In other microscopic fields, focal microscopic metallic debris may also be present within fatty marrow in the intertrabecular spaces of bone. For many periprosthetic tissue samples, nuclear detail is often obscured by intracellular debris, making recognition/differentiation of phagocytes difficult. As seen in Fig. 8, metallic debris is found intracellularly within

phagocytes (presumably macrophages). Fig. 9 shows plate-like corrosion products that are frequently observed in periprosthetic tissue samples from Prestige explants. As we have noted previously, this finding is consistent with histology from other stainless steel surgical devices.^{4, 35} Similarly, as shown in Fig. 10, hemosiderin is frequently found in periprosthetic tissue samples from Prestige explants. Hemosiderin (iron storage granules) may be present in areas of old hemorrhage or may be deposited in tissues with a high iron concentration. Histology of one of the explants has shown evidence of an infection.




Figure 7. Histology of periprosthetic tissues adjacent to a stainless steel Prestige cervical artificial disc showing intracellular and extracellular metallic debris in the interstices of fibrovascular connective tissues at the device interface (H&E stain, original magnification = 200×). (Color version of figure is available online.)




Figure 8. Histology of periprosthetic tissues adjacent to a stainless steel Prestige cervical artificial disc showing intracellular microscopic metallic debris within macrophages in fibrovascular tissues (Wright–Giemsa stain, original magnification = 500×). (Color version of figure is available online.)




Figure 9. Histology of periprosthetic tissues adjacent to a stainless steel Prestige cervical artificial disc showing the presence of plate-like corrosion products in fibrovascular connective tissues (H&E, original magnification = 400×). (Color version of figure is available online.)




Figure 10. Histology of periprosthetic tissues adjacent to a stainless steel Prestige cervical artificial disc showing hemosiderin (iron storage granules) in fibrovascular connective tissues (H&E, original magnification = 400×). (Color version of figure is available online.)

The typical host response found in periprosthetic tissue samples adjacent to the Prestige explants is characterized as a mononuclear chronic inflammatory response consisting of <15 mononuclear phagocytes (presumably macrophages) per 400× to 500× high-powered field. In addition, <5 neutrophils, lymphocytes, eosinophils, and foreign body giant cells per 400× to 500× high-powered fields are typically observed. Unlike the host response to Bryan explants (described later in detail), a multinuclear foreign body giant cell response is not commonly observed in periprosthetic tissues adjacent to Prestige devices (although a few foreign body giant cells may be found). ASTM F981-04 interprets <15 phagocytes in fields at 400× to 500× magnification as a score of 1 (mild reaction). The observed chronic inflammatory response represents a typical finding in periprosthetic tissues adjacent to the Prestige MOM spine arthroplasty devices. The observed host responses do not have an acute inflammatory character (few neutrophils). At this time, adverse events associated with degradation products, such as necrosis, osteolysis, or tissue degeneration, have not been observed in these explants. Neither lymphocyte infiltration nor accumulation of neutrophils has been observed in periprosthetic tissues.

Although stainless steel is well understood as a biomaterial, to date only a few intermediate-term artificial disc retrievals and tissues have been studied thus far with >5 years implantation time. Retrievals show evidence of wear on the explants and wear products in periprosthetic tissues. At present, we have not observed the adverse

local tissue reactions, previously reported with CoCr MOM bearings, for the stainless steel-on-stainless steel cervical bearings of the Prestige.

Polyurethanes and Titanium Alloy—Bryan Disc

The Bryan cervical artificial disc (Medtronic Spinal and Biologics) is a biarticular prosthesis fabricated from a pair of identical titanium alloy shells (which are fitted into the endplates of the cervical vertebral body), a PCU nucleus (which articulates with the titanium shells), and a flexible polyether urethane sheath (meant to prevent tissue ingrowth into the articulating surfaces), which surrounds the polyurethane nucleus. The Bryan cervical artificial disc is indicated for degenerative disc disease at one level from C3 to C7. Clinical studies of the Bryan started in Europe in 2000, and up to 8 years of clinical follow-up has been performed to date.^{40, 41}

We have analyzed 35 Bryan TDRs retrieved from 30 patients after 3.2 years (0.3-7.0 years). The Bryan retrievals were made available and funded through Medtronic's IDE studies or Medtronic's quality system. The nominal height loss of the explanted cores (mean \pm SD) was 0.44 ± 0.55 mm (range: 0.04-2.6 mm). We generally observed localized microscopic evidence of adhesive and abrasive wear (confirmed by SEM and interferometry), and attributed the majority of initial height loss to creep because, for many of the retrievals, the initial glossy surface finish of the cores was generally well preserved, even after 7.0 years in vivo (Fig. 11). We have observed endplate impingement in 9 of 30 (30%) of Bryan retrievals (Fig. 12). In 2 cases, the rim impingement generated Ti debris and caused third-body wear of the PCU core with substantial height loss (1.6 and 2.6 mm). In 1 case, impingement was associated with a full-thickness rim fracture of the PCU core.




Figure 11. Nucleus of a retrieved Bryan implant (6.1 years in vivo) showing its glossy appearance. (Color version of figure is available online.)




Figure 12. Endplate impingement in a retrieved Bryan implant (6.9y *in vivo*). (Color version of figure is available online.)

The sheaths typically showed evidence of folding, or permanent deformation in regions where the core made repeated contact. We also found evidence of *in vivo* degradation of the sheath in a subset of these devices (4/15, 27%), and the manifestation appears to be patient specific rather than associated with implantation time. In the early stages of biodegradation, the sheath gains an opaque or cloudy appearance (Fig. 13). Biodegradation of the polyether urethane sheath, in the later stages, leads to surface fissures and, in some cases, full-thickness cracks (Fig. 13). The mechanism thought to be responsible for biodegradation of the polyurethane sheath is caused by the release of reactive oxygen species by macrophages and foreign body giant cells.⁴² The clinical significance of sheath biodegradation appears to be minimal at the early stages in which the sheath becomes discolored, but if the sheath degrades *in vivo* over the long-term, it remains unclear whether the polymer will generate debris that contributes to an inflammatory reaction around the implant. None of our implants received to date have been revised owing to issues related to the sheath. We continue to collect Bryan retrievals to better understand long-term outcomes and potential complications.



Figure 13. Posterior view of the sheath of a retrieved Bryan implant (1.6y *in vivo*) showing the cloudy appearance along with areas of cracking. (Color version of figure is available online.)

We have analyzed periprosthetic tissue samples from approximately 15 Bryan cervical artificial discs that have been implanted between 1 and 6 years. Unlike MOM TDRs, few periprosthetic tissues show metallosis at gross examination. As seen in Figure 14, Figure 15, Figure 16, microscopically, foreign body giant cells and macrophages with intracellular polymeric particulate debris tend to be observed in periprosthetic tissue samples. Polarized light microscopy often reveals weakly birefringent intracellular polymeric debris in the histology of periprosthetic tissue samples (Fig. 14). Corrosion products are not frequently found in the histology of periprosthetic tissues. Previously published data from both preclinical animal studies and human clinical

explants have shown that the polyurethane polymers used in the sheath and core of the Bryan cervical disc are birefringent when viewed under polarized light microscopy and, thus, are readily recognized in histology sections.^{4, 35, 43, 44} As shown in Fig. 14, polarized light microscopy clearly demonstrates birefringent polymeric particulate in periprosthetic tissues. As a complement to polarized light microscopy, Oil Red-O staining, specific for the presence of polymeric wear debris in cleared tissues, can also be used to document the presence of polymeric debris. As shown in Figure 15, Figure 16, similar to polyethylene particulate in large joint histology, Oil Red-O staining is useful in demonstrating the presence of polyurethane polymeric debris (red) in periprosthetic tissue samples. Histology micrographs document that these complementary techniques reveal red-stained (when Oil Red-O sections were evaluated) birefringent polymeric debris in the tissues, most likely related to wear in the Bryan device. Histology techniques cannot determine whether the polymeric debris is from the sheath or nucleus of the Bryan device. No evidence of osteolysis has been observed in the histology associated with these explants. Histology of one of the explants has shown evidence of an infection.




Figure 14. Histology of periprosthetic tissues adjacent to a Bryan polyurethane cervical TDR showing birefringent intracellular polymeric particulate within macrophages (H&E, original magnification = 200x, partially polarized light). (Color version of figure is available online.)




Figure 15. Histology of periprosthetic tissues adjacent to a Bryan polyurethane cervical TDR showing larger red-stained extracellular polymeric particulate debris as well as foreign body giant cells and macrophages with smaller intracellular red-stained polymeric debris (Oil-Red-O, original magnification = 200x). (Color version of figure is available online.)




Figure 16. Histology of periprosthetic tissues adjacent to a Bryan cervical TDR showing intracellular red-stained polymeric debris within macrophages and foreign body giant cells (Oil Red-O stain with hematoxylin counterstain, original magnification = 500×). (Color version of figure is available online.)

Based on the population and type of inflammatory cells observed in the histology of these devices, a moderate-to-marked chronic inflammatory host response consisting of macrophages and foreign body giant cells is typically observed in the histology. In one of the explants, lobulated epithelioid granulomatous tissues, similar to pseudosynovial tissues found in large joint arthroplasty, are sometimes found in these periprosthetic tissues. In this case, focal lymphocyte infiltration with >20 lymphocytes was observed in 500× fields. Less than 5 eosinophils per 500× field were focally observed in these same fields for that explant. A sample histology image is seen in Fig. 17. As is the case with Bryan cervical TDRs, chronic inflammatory host responses with intracellular polymeric debris are frequently observed adjacent to joint arthroplasty devices such as TDRs.




Figure 17. Histology of periprosthetic tissues adjacent to a Bryan cervical TDR showing an inflammatory response consisting of macrophages, lymphocytes, and eosinophils (H&E stain, original magnification = 500×). (Color version of figure is available online.)

Case studies of osteolysis and adverse local tissue reactions have been observed with the Bryan artificial disc. In one case, osteolysis was attributed to a periprosthetic infection. The authors are aware of 2 reported cases of aseptic osteolysis associated with the Bryan; one of these cases has been presented at an international conference in Europe.⁴⁵ It concerned an asymptomatic young lady in whom the osteolysis of the vertebral bodies adjacent to the artificial disc was detected at a routine follow-up examination at 3 years postoperatively. The other case was a middle-aged man who developed a recurrent radiculopathy due to osteolysis and subsequent reactive hypertrophic bone formation with nerve root compression at 5 years postoperatively. In

both cases, the prosthesis was removed and an interbody fusion with autograft and plate stabilization performed. A full description of these cases is currently being prepared for a future journal publication.

The Role of Retrieval Analysis in the Development of New TDR Test Methods

The aforementioned retrieval studies provide standardization bodies with scientific data by which in vitro test methods for TDRs can and should be validated. At present, few standards for TDR characterization exist.^{46, 47, 48} The standards that do exist are focused on static and dynamic characterization and intended bearing wear of the devices. Unlike other arthroplasty devices, test methods for the other modes of wear, including impingement and third-body wear, remain under development.

Conveniently, retrieval analysis results provide guidance for the characterization techniques that can be used to correlate in vivo and in vitro wear patterns and morphology.⁴⁹ Several studies have succeeded in drawing correlations between the analyses conducted on retrieved and wear-tested devices.^{17, 35} These analyses focus on using optical microscopy, SEM, microcomputed tomography, and surface profilometry to quantify the characteristic wear mechanisms and damage patterns of the devices. There is now strong consensus that both the worst-case and most clinically relevant wear protocol for metal-on-polyethylene TDR bearing couples is one that uses a simultaneous multicomponent motion sequence that results in cross-shear stresses at the bearing interface.⁸ However, for MOM bearing couples, the worst-case and clinically relevant wear simulation for TDRs is likely not contained in one test. Rather, a unidirectional test typically presents a worst case for most MOM bearing couples, and yet, most clinical MOM TDR retrievals demonstrate multidirectional abrasive wear, indicating a simultaneous multicomponent motion sequence is more clinically relevant.

Although selecting the correct normal wear scenario for a given bearing couple is laden with trade-offs and the need for justification, when more adverse testing conditions are considered, even less guidance is available. As we have seen in the previous sections of this review, a substantial fraction of retrieved cervical and lumbar TDRs have been documented with evidence of in vivo impingement, regardless of design or bearing material. Using the available retrieval evidence, it is clear that TDR impingement is likely the result of a number of both design and clinical factors, including subsidence or suboptimal positioning.

The first step in developing a clinically relevant impingement wear testing approach begins with understanding the wear mechanisms and contact areas present on the retrieved devices. Next, understanding the ranges of motion and combined loading scenarios that generate the impingement scenario can be achieved virtually using modeling software or on the bench using range-of-motion testing procedures.^{50, 51} Finally, the loading scenarios should be replicated using a cyclic profile under lubricated conditions. The in vitro samples should be analyzed for wear rate, morphology, and mechanism. The fluid collected during testing should be analyzed for particle size distribution and morphology. All results should be compared with the normal or bearing wear test conducted on the same design and verified mechanistically against the retrieved devices. This approach is founded on the use of retrievals. As with the original TDR wear testing standards, as more artificial discs are evaluated, both material and design performance should become more predictable.

Standardization efforts are underway to increase guidance to the industry to assist with more thorough device characterization, including testing under adverse conditions. Developing a single set of boundary conditions that can be applied to all designs and material combinations is likely unachievable. Therefore, using the limited retrieval data available is perhaps the only scientifically valid approach to develop a set of justifiable testing conditions that mimic the in vivo performance of TDRs.

Summary and Conclusions

Our knowledge of TDR clinical performance based on explanted devices and tissues may no longer be in its infancy, but it still remains far from complete. The most retrieval data are available for historical TDRs in which polyethylene was effectively gamma irradiated in air. The wear and damage modes associated with these early devices conform to expectations of gamma-air-irradiated polyethylene from the literature. Contemporary TDRs incorporating conventional gamma-inert-sterilized polyethylene, approved by the FDA starting in 2004, exhibit an order of magnitude lower oxidation than the historical polyethylene devices based on the short-term retrievals that have been studied to date. Because of the markedly reduced oxidation, it is expected that conventional polyethylene in TDRs will exhibit lower risk of fatigue and fracture than the historical gamma-air-sterilized polyethylene TDRs. Although more advanced, highly crosslinked polyethylene technologies are currently used in the fabrication of hip and knee replacements, it remains unclear whether the improved wear resistance these new materials afford are adequately balanced by the increased fracture risk that accompanies elevated radiation crosslinking.

At present, many of the unanswered questions surrounding polyethylene in TDRs relate to better understanding the inflammatory tissue reactions to wear debris and how the inflammation can be effectively avoided, either by improved materials science (eg, radiation crosslinking of the polyethylene) or, potentially, by pharmacologic intervention. Furthermore, the relationships between TDR bearing design (eg, fixed vs. mobile bearing), wear debris' release, and impingement damage remain poorly understood. Because of the more complete clinical and retrieval history describing their performance, polyethylene TDRs provide the starting point for validating realistic wear and fatigue test protocols to characterize TDRs during the research and development phase of preclinical discovery. It is especially clear from the retrieval history that it is crucial to characterize TDR response under "normal" or expected use, as well as adverse conditions, including impingement, that are frequently observed in explanted devices. Regulatory agencies are considering making adverse testing and impingement characterization essential, rather than optional, components of a preclinical test program for new TDR designs seeking regulatory approval. The most effective and realistic impingement simulations will incorporate the specific biomechanical analysis and retrieval evidence, if available, to validate the device-specific conditions for the impingement testing.

Overall, the clinical data associated with TDR technology, regardless of bearing couple, suggest that device characterization and research efforts should be focused on validating the existing wear testing standards, using retrieval data to understand the clinical relevance of test duration (ie, the number of loading and motion cycles) and profile selection. The same retrieval data sets provide a basis for impingement protocol development and warrant the study of third-body debris generation. It remains unclear whether the effect of combined adverse conditions, such as impingement-related accelerated wear, will lead to third-body abrasive conditions at intended bearing surfaces. Further, the effect of all associated particle debris, whether from intended bearing surfaces or from impingement conditions, especially in nonpolymeric bearing couples, remains highly contentious and warrants investigation. Regardless of bearing material selection or design, the lessons learned and approaches to retrieval analysis and wear test validation used in hip and knee arthroplasty should be implemented for TDRs to assure patients, surgeons, and regulators of device safety and effectiveness.

Disclosure

The research studies described in this article were supported in part by funding from NIH R01 AR47904, NIH R01 AR56264, and the FDA Critical Path Program. Institutional support for the Maverick, Prestige, and Bryan retrieval programs was received from Medtronic-funded IDE studies and Medtronic's quality system. The institutions of one or more of the authors have received institutional funding from Medtronic, Stryker, Zimmer, Biomet,

Depuy, InVivo, and Ticon. Dr. Toth is a scientific consultant to Medtronic, however he has not received any remuneration as a consultant in the past 12 months.

Acknowledgments

The authors thank Genevieve Hill and Jonathan Peck, FDA, for their helpful discussions and collaboration on studying impingement in TDR, as well as Frank Chan, Megan Harper, Jacob Shorez, and Mark Dace, Medtronic, for their contributions. The authors gratefully acknowledge Amy Rizzo, MT(ASCP), for her extraordinary skill and technical expertise in producing histological sections of periprosthetic tissues.

References

- 1 S.M. Kurtz. **The UHMWPE Biomaterials Handbook: Ultra-High Molecular Weight Polyethylene in Total Joint Replacement and Medical Devices.** ((ed 2)), Academic Press, Burlington, MA (2009)
- 2 S.M. Kurtz. **Total disc arthroplasty.** S.M. Kurtz, A.A. Edidin (Eds.), *Spine Technology Handbook*, Academic Press, New York, NY (2006), p. 303
- 3 S.M. Kurtz, M.J. Steinbeck, A. Ianuzzi, *et al.* **Retrieval analysis of motion preserving spinal devices and periprosthetic tissues.** *SAS J*, 3 (2009), pp. 161-177
- 4 P.A. Anderson, S.M. Kurtz, J.M. Toth. **Explant analysis of total disc replacement.** *Semin Spine Surg*, 18 (2006), pp. 1-12
- 5 S.M. Kurtz, H.A. Gawel, J.D. Patel. **History and systematic review of wear and osteolysis outcomes for first-generation highly crosslinked polyethylene.** *Clin Orthop Relat Res*, 469 (2011), pp. 2262-2277
- 6 K.J. Bozic, S. Kurtz, E. Lau, *et al.* **The epidemiology of bearing surface usage in total hip arthroplasty in the United States.** *J Bone Joint Surg*, 91 (2009), pp. 1614-1620
- 7 K.J. Bozic, K. Ong, E. Lau, *et al.* **Risk of complication and revision total hip arthroplasty among medicare patients with different bearing surfaces.** *Clin Orthop Relat Res*, 468 (2010), pp. 2357-2362
- 8 S. Kurtz, R. Siskey, L. Ciccarelli, *et al.* **Retrieval analysis of total disc replacements: implications for standardized wear testing.** *J ASTM Int*, 3 (2006), pp. 1-12
- 9 S.M. Kurtz, D. Macdonald, A. Ianuzzi, *et al.* **The natural history of polyethylene oxidation in total disc replacement.** *Spine*, 34 (2009), pp. 2369-2377
- 10 L. Costa, P. Bracco, E.M. Brach del Prever, *et al.* **Oxidation and oxidation potential in contemporary packaging for polyethylene total joint replacement components.** *J Biomed Mater Res A*, 78 (2006), pp. 20-26
- 11 D.A. Baker, A. Bellare, L. Pruitt. **The effects of degree of crosslinking on the fatigue crack initiation and propagation resistance of orthopedic-grade polyethylene.** *J Biomed Mater Res A*, 66 (2003), pp. 146-154
- 12 H.D. Link, A. Keller. **Biomechanics of total disc replacement.** K. Buttner-Janz, S.H. Hochschuler, P.C. McAfee (Eds.), *The Artificial Disc*, Springer, Berlin (2003), pp. 33-52
- 13 A. van Ooij, F.C. Oner, A.J. Verbout. **Complications of artificial disc replacement: a report of 27 patients with the SB Charité disc.** *J Spinal Disord Tech*, 16 (2003), pp. 369-383
- 14 A. van Ooij, S.M. Kurtz, F. Stessels, *et al.* **Polyethylene wear debris and long-term clinical failure of the Charité disc prosthesis: a study of 4 patients.** *Spine*, 32 (2007), pp. 223-229
- 15 I.M. Punt, V.M. Visser, L.W. van Rhijn, *et al.* **Complications and reoperations of the SB Charité lumbar disc prosthesis: experience in 75 patients.** *Eur Spine J*, 17 (2008), pp. 36-43
- 16 S.M. Kurtz, A. van Ooij, R. Ross, *et al.* **Polyethylene wear and rim fracture in total disc arthroplasty.** *Spine J*, 7 (2007), pp. 12-21
- 17 S.M. Kurtz, A. Patwardhan, D. MacDonald, *et al.* **What is the correlation of in vivo wear and damage patterns with in vitro TDR motion response?** *Spine*, 33 (2008), pp. 481-489
- 18 I.M. Punt, S. Austen, J.P. Cleutjens, *et al.* **Are periprosthetic tissue reactions observed after revision of total disc replacement comparable to the reactions observed after total hip or knee revision surgery?** *Spine (Phila Pa 1976)*, 37 (2012), pp. 150-159

- 19 I. Punt, R. Baxter, A. van Ooij, *et al.* **Submicron sized ultrahigh molecular weight polyethylene wear particle analysis from revised SB Charite III total disc replacements.** *Acta Biomater*, 7 (2011), pp. 3404-3411
- 20 I.M. Punt, J.P. Cleutjens, T. de Bruin, *et al.* **Periprosthetic tissue reactions observed at revision of total intervertebral disc arthroplasty.** *Biomaterials*, 30 (2009), pp. 2079-2084
- 21 D.R. Lebl, F.P. Cammisa, F.P. Girardi, *et al.* **Analysis of wear, surface properties, and fixation of retrieved lumbar total disc replacements: Proceedings of the NASS 26th Annual Meeting.** (2011)
- 22 N. Kelly, J. Cottrell, T. Wright. **Retrieval analysis of ProDisc total disc replacements, in Transactions of the 54th Meeting for the Orthopaedic Research Society.** (2008)
- 23 D.R. Lebl, F.P. Cammisa, F.P. Girardi, *et al.* **Analysis of wear, surface properties, and fixation of explanted cervical total disc replacements, in Proceedings of the NASS 26th Annual Meeting.** (2011)
- 24 T. Wright, J. Cottrell. **Retrieval analysis of ProDisc total disc replacements.** Presented at the *7th Annual Meeting of the Spine Arthroplasty Society*, May 1-4, Berlin, Germany (2007), p. 124
- 25 W. Kafer, C.B. Clessienne, M. Daxle, *et al.* **Posterior component impingement after lumbar total disc replacement: a radiographic analysis of 66 ProDisc-L prostheses in 56 patients.** *Spine* (Phila, PA 1976), 33 (2008), pp. 2444-2449
- 26 T.J. Choma, J. Miranda, R. Siskey, *et al.* **Retrieval analysis of a prodisc-L total disc replacement.** *J Spinal Disord Tech*, 22 (2009), pp. 290-296
- 27 J.J. Jacobs, R.M. Urban, N.J. Hallab, *et al.* **Metal-on-metal bearing surfaces.** *J Am Acad Orthop Surg*, 17 (2009), pp. 69-76
- 28 D.J. Langton, S.S. Jameson, T.J. Joyce, *et al.* **Early failure of metal-on-metal bearings in hip resurfacing and large-diameter total hip replacement: a consequence of excess wear.** *J Bone Joint Surg Br*, 92 (2010), pp. 38-46
- 29 S. Glyn-Jones, H. Pandit, Y.M. Kwon, *et al.* **Risk factors for inflammatory pseudotumour formation following hip resurfacing.** *J Bone Joint Surg Br*, 91 (2009), pp. 1566-1574
- 30 Food and Drug Administration. **Metal-on-metal hip implants.** <http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/ImplantsandProsthetics/MetalonMetalHipImplants/default.htm>. Accessed August 26, 2011
- 31 **Medical device alert: All metal-on-metal (MoM) hip replacements.** *Medicines and Healthcare Products Regulatory Agency*, Department of Health. 1, London (2010)
- 32 B. Meier. **In medicine, new isn't always improved.** *The New York Times* (2011)
- 33 B. Meier. **Hip implant complaints surge, even as the dangers are studied.** *The New York Times* (2011)
- 34 M.A. Wimmer, A. Fischer, R. Büscher, *et al.* **Wear mechanisms in metal-on-metal bearings: the importance of tribochemical reaction layers.** *J Orthop Res*, 28 (2010), pp. 436-443
- 35 P.A. Anderson, J.P. Rouleau, J.M. Toth, *et al.* **A comparison of simulator-tested and -retrieved cervical disc prostheses: Invited submission from the joint section meeting on disorders of the spine and peripheral nerves, March 2004.** *J Neurosurg*, 1 (2004), pp. 202-210
- 36 R.D. Guyer, J. Shellock, B. MacLennan, *et al.* **Early failure of metal-on-metal artificial disc prostheses associated with lymphocytic reaction: diagnosis and treatment experience in four cases.** *Spine*, 36 (2011), pp. E492-E497
- 37 D.A. Cavanaugh, P.D. Nunley, E.J. Kerr 3rd, *et al.* **Delayed hyper-reactivity to metal ions after cervical disc arthroplasty: a case report and literature review.** *Spine*, 34 (2009), pp. E262-E265
- 38 M.R. Berry, B.G. Peterson, D.H. Alander. **A granulomatous mass surrounding a maverick total disc replacement causing iliac vein occlusion and spinal stenosis: a case report.** *J Bone Joint Surg Am*, 92 (2010), pp. 1242-1245
- 39 L.M. Tumialán, W.M. Gluf. **Progressive vertebral body osteolysis after cervical disc arthroplasty.** *Spine*, 36 (2011), pp. E973-E978
- 40 J. Walraevens, P. Demaerel, P. Suetens, *et al.* **Longitudinal prospective long-term radiographic follow-up after treatment of single-level cervical disc disease with the Bryan cervical disc.** *Neurosurgery*, 67 (2010) 679-387

- 41 J. Goffin, J. van Loon, F. Van Calenbergh, *et al.* **A clinical analysis of 4- and 6-year follow-up results after cervical disc replacement surgery using the Bryan cervical disc prosthesis.** *J Neurosurg Spine*, 12 (2010), pp. 261-269
- 42 E.M. Christenson, M. Dadsetan, M. Wiggins, *et al.* **Poly(carbonate urethane) and poly(ether urethane) biodegradation: in vivo studies.** *J Biomed Mater Res A*, 69 (2004), pp. 407-416
- 43 P.A. Anderson, J.P. Rouleau, V.E. Bryan, *et al.* **Wear analysis of the Bryan cervical disc prosthesis.** *Spine*, 28 (2003), pp. S186-S194
- 44 P.A. Anderson, R.C. Sasso, J.P. Rouleau, *et al.* **The Bryan cervical disc: wear properties and early clinical results.** *Spine J*, 4 (Suppl 6) (2004), pp. S303-S309
- 45 J. Goffin, J. van Loon, F. Van Calenbergh, *et al.* **Place of cervical artificial disc implantations.** *Transactions of the European Association of Neurosurgical Societies (EANS)*, Groningen, The Netherlands (2010)
- 46 **ASTM F2423 standard guide for functional, kinematic, and wear assessment of total disc prostheses.** (2011)
- 47 **ISO 18192-1: Implants for surgery—Wear of total intervertebral spinal disc prostheses—Part 1: loading and displacement parameters for wear testing and corresponding environmental conditions for test.** (2008)
- 48 **ASTM F2346 Standard Test Methods for Static and Dynamic Characterization of Spinal Artificial Discs.** (2005)
- 49 **ASTM F561-05a 2010 Standard Practice for Retrieval and Analysis of Medical Devices, and Associated Tissues and Fluids.** (2010)
- 50 D.J. Rundell Sa, J. Isaza, R. Siskey, *et al.* **Derivation of clinically relevant boundary conditions suitable for evaluation of chronic impingement of lumbar total disc replacement: application to standard development.** *Int J ASTM*, 8 (2011), pp. 1-14
- 51 H.J. Wilke, K. Wenger, L. Claes. **Testing criteria for spinal implants: recommendations for the standardization of *in vitro* stability testing of spinal implants.** *Eur Spine J*, 7 (1998), pp. 148-154