Marquette University

e-Publications@Marquette

Mathematics, Statistics and Computer Science Faculty Research and Publications Mathematics, Statistics and Computer Science, Department of (- 2019)

2-2017

Incorporating a Spatial Prior into Nonlinear D-Bar EIT Imaging for Complex Admittivities

Sarah J. Hamilton

Marquette University, sarah.hamilton@marquette.edu

Jennifer L. Mueller

Colorado State University - Fort Collins

Melody Alsaker

Colorado State University - Fort Collins

Follow this and additional works at: https://epublications.marquette.edu/mscs_fac

Part of the [Computer Sciences Commons](#), [Mathematics Commons](#), and the [Statistics and Probability Commons](#)

Recommended Citation

Hamilton, Sarah J.; Mueller, Jennifer L.; and Alsaker, Melody, "Incorporating a Spatial Prior into Nonlinear D-Bar EIT Imaging for Complex Admittivities" (2017). *Mathematics, Statistics and Computer Science Faculty Research and Publications*. 471.

https://epublications.marquette.edu/mscs_fac/471

Marquette University

e-Publications@Marquette

Mathematics, Statistics and Computer Sciences Faculty Research and Publications/College of Arts and Sciences

This paper is NOT THE PUBLISHED VERSION; but the author's final, peer-reviewed manuscript. The published version may be accessed by following the link in the citation below.

IEEE Transactions on Medical Imaging, Vol. 36, No. 2 (2017): 457-466. [DOI](#). This article is © Institute of Electrical and Electronic Engineers (IEEE) and permission has been granted for this version to appear in [e-Publications@Marquette](#). Institute of Electrical and Electronic Engineers (IEEE) does not grant permission for this article to be further copied/distributed or hosted elsewhere without the express permission from Institute of Electrical and Electronic Engineers (IEEE).

Incorporating a Spatial Prior into Nonlinear D-Bar EIT Imaging for Complex Admittivities

Sarah J. Hamilton

Department of Mathematics Statistics, and Computer Science, Marquette University, Milwaukee, WI

J. L. Mueller

Department of Mathematics and School of Biomedical Engineering, Colorado State University, Fort Collins, CO

M. Alsaker

Department of Mathematics, Colorado State University, Fort Collins, CO

Abstract:

Electrical Impedance Tomography (EIT) aims to recover the internal conductivity and permittivity distributions of a body from electrical measurements taken on electrodes on the surface of the body. The reconstruction task is a severely ill-posed nonlinear inverse problem that is highly sensitive to measurement noise and modeling errors. Regularized D-bar methods have shown great promise in producing noise-robust algorithms by employing a low-pass filtering of nonlinear (nonphysical) Fourier transform data specific to the EIT problem. Including prior data with the approximate locations of major organ boundaries in the scattering transform provides a means of extending the radius of the low-pass filter to include higher frequency components in the

reconstruction, in particular, features that are known with high confidence. This information is additionally included in the system of D-bar equations with an independent regularization parameter from that of the extended scattering transform. In this paper, this approach is used in the 2-D D-bar method for admittivity (conductivity as well as permittivity) EIT imaging. Noise-robust reconstructions are presented for simulated EIT data on chest-shaped phantoms with a simulated pneumothorax and pleural effusion. No assumption of the pathology is used in the construction of the prior, yet the method still produces significant enhancements of the underlying pathology (pneumothorax or pleural effusion) even in the presence of strong noise.

SECTION I. Introduction

Electrical Impedance Tomography (EIT) is a non-invasive radiation-free imaging modality in which low amplitude current is applied through electrodes placed on the surface of a body and the resulting voltages are measured. From these surface measurements, images of the interior conductivity and permittivity can be obtained. The severe ill-posedness of the inverse conductivity/permittivity problem limits the spatial resolution of the reconstructed images, which hinders their clinical applicability. Mathematical stability estimates for the problem can be found in,^{1-2,3,4} for example. The use of spatial *a priori* information in the solution of the inverse problem provides a means of including anatomical information that is present with high confidence, while still allowing unknown features such as lung pathologies to emerge in the reconstructed image without any assumption of their presence. In patients with serious respiratory illness, it is often the case that a CT scan is performed to obtain a diagnosis or for a regular exam in the case of a chronic illness, and the condition is monitored with one or more follow-up scans. The initial scan can provide basic *a priori* information for the reconstruction algorithm such as chest shape, approximate lung and heart sizes, and relative positions in the plane of the electrodes.

The state-of-the-art of EIT in critical care thoracic imaging is to image 2-D slices in the plane of a belt of electrodes placed around the thorax.^{5-6,7,8,9,10} Estimates of alveolar collapse and compliance can be determined from reconstructions of impedance changes in lung regions in these slices, and the gravitational vector inside the lung can be taken into account.¹¹ Ventilation and perfusion maps can also be created in these slices in real time, providing a method of monitoring PEEP titration.⁷

A priori information has been used successfully in iterative reconstruction algorithms to enhance image quality (see, for example,^{12-13,14,15,16,17,18,19,20,21} for a partial list), and more recently in²² in the direct 2-D D-bar method for (real-valued) conductivity reconstruction. In this paper, the method of²² is extended to the 2-D D-bar algorithm for the reconstruction of complex admittivities.^{23-24,25} The reconstruction algorithm for complex admittivities differs from the D-bar algorithm for real-valued conductivities in the construction of the complex geometrical optics (CGO) solutions. While the well-developed real-valued case^{26-27,28,29,30} utilizes the familiar transformation of the generalized Laplace equation governing the physical EIT problem to a Schrödinger equation, the complex admittivity algorithm requires transforming the problem to a first order elliptic system and constructing two sets of CGO solutions. The algorithm is described briefly in Section 2, and the reader is referred to^{23-24, 25} for further detail. In²² a method of including a priori information in the 2-D D-bar algorithm for real-valued conductivities was proposed and proved to be a nonlinear regularization strategy. It was tested on simulated data on a circular domain. The 2-D D-bar algorithm for reconstructing complex admittivities differs from the algorithm in²² in several ways. First,

a 2×2 elliptic system of equations must be solved to recover the complex admittivity, and so the scattering transform becomes a scattering matrix, and a system of first order PDEs in ∂_z and $\partial_{\bar{z}}$ arises for the CGO solutions. In contrast to the method in,²² there are, in fact, two sets of CGO solutions to compute (described in Section II-B). Second, in contrast to,²² determining the admittivity requires an additional computational step beyond evaluation of the CGO solution at complex frequency $k = 0$, as is the case for the methods based on.³¹ In this work, the *a priori* information is included in the high-frequency component of the scattering matrix and in the appropriate CGO solution computations, which consist of two decoupled systems of integral equations. Thus, while the *a priori* information is in the analogous functions to those in,²² the overall algorithm contains some significant differences. The algorithm presented here is tested on realistic numerical phantoms on noncircular domains simulating a cross-sectional human chest. Similar phantoms have been used in, for example.^{14, 32, and 33}

The method incorporates spatial *a priori* information about the admittivity distribution in the scattering transform, as well as in the system of D-bar equations, and includes regularization parameters in each place that can be adjusted to control the amount of influence the prior has on the reconstruction. This approach does come at the cost of solving the forward problem once using the prior admittivity distribution. For the D-bar method, the forward problem that is solved using the prior is not the generalized Laplace equation with the complete electrode model that is typically solved at each iteration by, for example, the finite element method, but rather the equations that define the CGO solutions. Also, depending upon how the prior is constructed, a preliminary reconstruction (that could be computed using the D-bar method with no prior, or a different method) may be required. In contrast, iterative methods include the prior in the cost functional in the regularization term, which does not come at additional computational cost. Both types of method run the risk of biasing the solution with the *a priori* information, and care must be taken in weighting such information. Depending upon the choice of weighting parameters, the segmented prior can be very strong or very weak. Post-processing methods also run the risk of introducing bias in the images.

The effectiveness of the method is tested here on simulated data with 0.1% and 1.0% added Gaussian relative noise for a 2-D phantom chest with a simulated pleural effusion and with a simulated pneumothorax. No *a priori* information about the presence of the effusion or the pneumothorax is used in the reconstruction, only *a priori* spatial information about the heart and lung boundaries. Nevertheless, both the effusion and pneumothorax become considerably sharper than in images computed without the *a priori* organ boundary information.

The organ boundaries of the “heart and lungs prior” are depicted in Figure 1(a). While the initial prior is piecewise constant, after conductivity and permittivity values have been assigned, the prior is mollified to obtain a smooth function since the method of computing the scattering transform for the prior requires that it be differentiated. Assigning the initial admittivity values to the prior can be done in a number of ways, and the *a priori* reconstruction algorithm presented here is valid for any assignment method. In our tests, we computed an initial reconstruction with no prior from the noisy data (which we will refer to as a standard D-bar reconstruction), then computed the maximum conductivity/permittivity in the heart region, and minimum conductivity/permittivity in each lung region of the piecewise constant “heart and lungs

prior”, and assigned those values to each respective region of the piecewise constant prior. Further implementation details are found in Section III-B.

Fig. 1. The “heart and lungs” phantom (a) and the test examples studied to simulate two pathologies: (b) a pneumothorax in the ventral part of the left lung and (c) a pleural effusion in the dorsal part of the left lung. The black lines correspond to true boundaries in the simulations, and the superimposed red dots are the organ boundaries used in the construction of the admittivity prior γ_{PR} before smoothing.

The paper is organized as follows. The *a priori* method is presented in Section II, which first provides a brief description of the forward model in Section II-A used to simulate the EIT data, followed by a summary of the D-bar method for complex admittivity imaging in Section II-B, with the modifications for the *a priori* method described in Section II-C. The D-bar method for admittivity reconstruction is admittedly mathematically complicated, and the reader is referred to the papers ^{24, 25, 34} for further details. The numerical implementation and testing of the method is outlined in Section III, and the discussion and conclusions presented in Section IV.

SECTION II. Methods

A. The Forward Model

The electric potential $u(x, y)$ inside the 2-D region Ω is modeled by the *admittivity equation*, a generalized Laplace equation,

$$\nabla \cdot \gamma(x, y) \nabla u(x, y) = 0, \quad (1)$$

(1)

where $\gamma(x, y) = \sigma(x, y) + i\omega\epsilon(x, y)$ denotes the complex valued admittivity, $\sigma(x, y)$ the electrical conductivity (bounded away from zero $0 < \sigma(x, y) < C$), $\epsilon(x, y)$ the electrical permittivity (assumed to be non-negative), and ω the angular frequency of the applied current. The boundary data for the inverse problem is the *Dirichlet-to-Neumann* (DN) map Λ_γ which maps a voltage at the boundary to the

corresponding current density, i.e., $\Lambda_\gamma: u|_{\partial\Omega} \mapsto \gamma \frac{\partial u}{\partial \nu}|_{\partial\Omega}$, where ν denotes the outward unit normal vector to the boundary $\partial\Omega$. In practice, to dampen rather than amplify the noise in the measured data, currents are applied and the resulting voltages are measured. This corresponds to knowledge of the *Neumann-to-Dirichlet* ND map $R_\gamma: \gamma \frac{\partial u}{\partial \nu}|_{\partial\Omega} \mapsto u|_{\partial\Omega}$. Ensuring conservation of charge and specifying a ground, the ND map can be inverted to obtain the DN map $\Lambda_\gamma = (R_\gamma)^{-1}$.

For the simulation of the data, a finite element implementation of the complete electrode model (CEM) was used. The CEM³⁵ takes into account both the shunting effect of the electrodes and the contact impedances between the electrodes and tissue. The complete electrode model consists of the admittivity [equation \(1\)](#) and the following boundary conditions on L electrodes:

$$\begin{aligned} u + z_l \gamma \frac{\partial u}{\partial \nu} &= U_l, x \in e_l, l = 1, 2, \dots, L \\ \int_{e_l} \gamma \frac{\partial u}{\partial \nu} dS &= J_l, x \in e_l, l = 1, 2, \dots, L, \\ \gamma \frac{\partial u}{\partial \nu} &= 0, x \in \partial\Omega \setminus \cup_{l=1}^L e_l, \end{aligned}$$

where z_l is the effective contact impedance between the l^{th} electrode e_l and the medium, J_l is the applied current, and U_l is the measured voltage. In addition, Kirchhoff's Law and the choice of ground must be imposed to ensure existence and uniqueness of the result:

$$\sum_{l=1}^L J_l = 0, \text{ and } \sum_{l=1}^L U_l = 0.$$

The uniqueness and existence of a solution to the CEM has been proven in.³⁶

Since the D-bar method, described below, relies on CGO solutions to equations that do not include the CEM, there is an inherent modeling error introduced in the reconstructions.

B. The D-Bar Method For Complex Admittivities

D-bar methods are named for the partial derivatives with respect to the complex conjugates that arise in the equations in the methods. The $\bar{\partial}$ operator with respect to the complex variable $z = x + iy$ and the related operator ∂_z are defined by

$$\bar{\partial}_z = \frac{1}{2} \left(\frac{\partial}{\partial x} + i \frac{\partial}{\partial y} \right), \partial_z = \frac{1}{2} \left(\frac{\partial}{\partial x} - i \frac{\partial}{\partial y} \right).$$

Throughout the paper, $\mathbb{R}^2$ is associated with $\mathbb{C}$ via $z = (x, y) \mapsto x + iy$.

The method described below is based on the uniqueness proof for the inverse admittivity problem,³⁷ which was completed as a constructive proof in.^{23, 24} The work³⁷ was based on the real-valued paper [38] in which a 2×2 elliptic system was introduced to reduce the smoothness required in the constructive uniqueness proof.³¹ An implementation of³⁸ was described in.^{39 and 40}

With the introduction of a non-physical complex parameter k , the admittivity [equation \(1\)](#) admits solutions with special exponentially growing behavior known as CGO solutions. In particular, it was shown in ²⁴ that there exist separate solutions $u_1(z, k)$ and $u_2(z, k)$ to [\(1\)](#) such that $u_1(z, k) \sim \frac{e^{ikz}}{ik}$ and $u_2(z, k) \sim -\frac{e^{-ik\bar{z}}}{ik}$ for large $|k|$ or $|z|$.

Defining an operator vector $\mathcal{D} = \gamma^{1/2}(\partial_z, \bar{\partial}_z)^T$, the change of variables

$$Q(z) = \begin{bmatrix} 0 & -\frac{1}{2}\partial_z \log \gamma(z) \\ -\frac{1}{2}\bar{\partial}_z \log \gamma(z) & 0 \end{bmatrix}, \quad (2)$$

and

$$(M_{11}, M_{21})^T = e^{-ikz} \mathcal{D}u_1, \quad (M_{12}, M_{22})^T = e^{ik\bar{z}} \mathcal{D}u_2 \quad (3)$$

transform the admittivity [equation \(1\)](#) into the first order elliptic system ³⁷

$$D_k M(z, k) - Q(z)M(z, k) = 0, \quad (4)$$

where

$$D_k M(z, k) = \begin{bmatrix} \bar{\partial}_z & 0 \\ 0 & \partial_z \end{bmatrix} M - ik \begin{bmatrix} 1 & 0 \\ 0 & -1 \end{bmatrix} \begin{bmatrix} 0 & M_{12} \\ M_{21} & 0 \end{bmatrix}.$$

[Equation \(4\)](#) has a unique solution $M(\cdot, k)$ where each component of $M(\cdot, k) - I$ is in $L^p(\mathbb{R}^2)$ for some $p > 2$, where I denotes the 2×2 identity matrix.

D-bar methods follow the basic computational outline:

$$\text{DN map} \mapsto \begin{matrix} \text{Scattering} \\ \text{Data} \end{matrix} \mapsto \begin{matrix} \text{CGO} \\ \text{Solutions} \end{matrix} \mapsto \text{Admittivity}.$$

The scattering data is a 2×2 matrix function $S(k)$, not directly physically measurable from the data, with zero entries on the diagonal and off-diagonal entries given by

$$\begin{aligned} S_{12}(k) &= \frac{i}{\pi} \int_{\Omega} Q_{12}(z) e(z, -\bar{k}) M_{22}(z, k) dx dy \\ S_{21}(k) &= -\frac{i}{\pi} \int_{\Omega} Q_{21}(z) e(z, k) M_{11}(z, k) dx dy \end{aligned} \quad (5)$$

where $e(z, k) \equiv e^{i(kz + \bar{k}\bar{z})}$ and $\text{supp } Q(z) \subseteq \bar{\Omega}$ from [\(2\)](#).

The DN map Λ_γ uniquely determines the scattering data $S(k)$, and the scattering data uniquely determines the admittivity $\gamma(z)$.³⁷ However, the relationship between the scattering data and the DN map relies on the intermediate computation of the CGO solutions u_1 and u_2 on the boundary of Ω as well as functions $\Psi_{12}(z, k) \equiv e^{-ik\bar{z}} M_{12}(z, k)$ and $\Psi_{21}(z, k) \equiv e^{ikz} M_{21}(z, k)$. This is described in Step 1 below.

Step 1: From Boundary Measurements to Scattering Data:

For each $|k| \leq R$, solve the following two boundary integral equations

$$\begin{aligned} u_1(z, k) &= \frac{e^{ikz}}{ik} - \int_{\partial\Omega} G_k(z - \zeta) (\delta\Lambda_\gamma) u_1(\zeta, k) ds(\zeta) \\ u_2(z, k) &= -\frac{e^{-ik\bar{z}}}{ik} - \int_{\partial\Omega} G_k(\bar{\zeta} - \bar{z}) (\delta\Lambda_\gamma) u_2(\zeta, k) ds(\zeta) \end{aligned} \quad (6)$$

for the traces of the CGO solutions u_1 and u_2 on the boundary. Here $G_k(z)$ denotes the Faddeev Green's function for the Laplace operator given by (see^{41, 42}),

$$G_k(z) = e^{ikz} \int_{\mathbb{R}^2} \frac{e^{iz \cdot \xi}}{\xi(\bar{\xi} + 2k)} d\xi,$$

and $\delta\Lambda_\gamma = \Lambda_\gamma - \Lambda_1$ where Λ_1 denotes the DN map corresponding to a constant admittivity 1 and $\gamma = 1$ near $\partial\Omega$. Provided that γ is a constant γ_b near $\partial\Omega$, the DN map can be scaled as in.²⁸

Next, compute the traces of the CGO solutions Ψ_{12} and Ψ_{21} from the second set of boundary integral equations

$$\begin{aligned} \Psi_{12}(z, k) &= p.v. \int_{\partial\Omega} \frac{e^{i\bar{k}(z-\zeta)}}{4\pi(z-\zeta)} (\delta\Lambda_\gamma) u_2(\zeta, k) ds(\zeta) \\ \Psi_{21}(z, k) &= p.v. \int_{\partial\Omega} \overline{\left[\frac{e^{ik(z-\zeta)}}{4\pi(z-\zeta)} \right]} (\delta\Lambda_\gamma) u_1(\zeta, k) ds(\zeta), \end{aligned} \quad (7)$$

where $p.v.$ denotes the principal value of the integral.

Then, compute the scattering transforms $S_{12}(k)$ and $S_{21}(k)$:

$$\begin{aligned} S_{12}(z, k) &= \frac{i}{2\pi} \int_{\partial\Omega} e^{-i\bar{k}z} \Psi_{12}(z, k) v(z) ds(z) \\ S_{21}(z, k) &= -\frac{i}{2\pi} \int_{\partial\Omega} e^{i\bar{k}\bar{z}} \Psi_{21}(z, k) \overline{v(z)} ds(z), \end{aligned} \quad (8)$$

where $v(z) = (v_1(z), v_2(z)) = v_1(z) + iv_2(z)$ denotes the outward unit normal to $\partial\Omega$ at the point z , and $\overline{v(z)}$ denotes its complex conjugate. The parameter R acts as a low-pass cutoff radius of the scattering data used to stabilize the reconstruction method in the presence of noise. The scattering data is set to zero

for $|k| > R$. This approach has been proved to be a nonlinear regularization strategy in the case of real-valued conductivities.⁴³ Parallel computing can be used to solve equations (6) – (8) since each of these equations is solved for each k independently. Further implementation details are found in Section III-B.

Step 2: Computation of CGO Solutions:

Let $\overline{\Omega}^+$ be a domain slightly larger than Ω . This will be needed to numerically compute the ∂_z and $\overline{\partial}_z$ derivatives of the CGO solutions $M(z, 0)$ required to form the matrix potential $Q(z)$ in Step 3. For each $z \in \overline{\Omega}^+$, solve the $\overline{\partial}_k$ equation

$$\overline{\partial}_k M(z, k) = M(z, \overline{k}) \begin{bmatrix} e(z, \overline{k}) & 0 \\ 0 & e(z, -k) \end{bmatrix} S(k), \quad (9)$$

using the fundamental solution $\frac{1}{\pi k}$ for the $\overline{\partial}_k$ operator, by solving the decoupled systems

$$\begin{cases} M_{11}(z, k) = 1 + \frac{1}{\pi k} * [M_{12}(z, \overline{k})e(z, -k)S_{21}(k)] \\ M_{12}(z, k) = 0 + \frac{1}{\pi k} * [M_{11}(z, \overline{k})e(z, \overline{k})S_{12}(k)] \end{cases} \quad (10)$$

$$\begin{cases} M_{21}(z, k) = 0 + \frac{1}{\pi k} * [M_{22}(z, \overline{k})e(z, -k)S_{21}(k)] \\ M_{22}(z, k) = 1 + \frac{1}{\pi k} * [M_{21}(z, \overline{k})e(z, \overline{k})S_{12}(k)]. \end{cases} \quad (11)$$

The convolutions $*$ take place in k over the disc of radius R as $S_{ij}(k)$ now has compact support. The systems are solved numerically as in.⁴⁴

Step 3: From CGO Solutions to the Admittivity:

Using the CGO solutions corresponding to $k = 0$, compute the potentials (only one is actually needed)

$$\begin{aligned} Q_{12}(z) &= \frac{\partial_z [M_{11}(z, 0) + M_{12}(z, 0)]}{M_{22}(z, 0) + M_{21}(z, 0)} \\ Q_{21}(z) &= \frac{\overline{\partial}_z [M_{22}(z, 0) + M_{21}(z, 0)]}{M_{11}(z, 0) + M_{12}(z, 0)}, \end{aligned} \quad (12)$$

and from these, compute the admittivity $\gamma(z)$ using either

$$\gamma(z) = \exp \left\{ -\frac{2}{\pi \overline{z}} * Q_{12}(z) \right\} = \exp \left\{ -\frac{2}{\pi z} * Q_{21}(z) \right\}, \quad (13)$$

where the convolution in z takes place over $\overline{\Omega}$ since Q has compact support.

The reader is referred to^{23, 25, 34} for developments and implementations of numerical algorithms for complex D-bar EIT imaging.

C. Inclusion of a priori Admittivity Information

The low-pass filtering (setting $S(k) = 0$ for $|k| > R$) in the non-physical scattering domain has an effect similar to that of traditional low-pass filtering in the standard Fourier domain. As $|k| \rightarrow \infty$, $M(\cdot, k) \sim I$ and thus the scattering data in (5) reduces to $S_{12}(k) \approx \frac{i}{\pi} \hat{Q}_{12}(2k_1, 2k_2)$ and $S_{21}(k) \approx -\frac{i}{\pi} \hat{Q}_{21}(-2k_1, 2k_2)$. Thus, for large $|k|$ the scattering data are essentially shifted Fourier transforms of the potential $Q(z)$. Hence, it is reasonable to expect a loss of sharp edges in reconstructions of $\gamma(z)$ from the low-pass filtered scattering data.

In practice, the scattering data computed via the boundary integral equations (8) “blows up” in magnitude as $|k|$ increases, sometimes as early as $|k| = 3.5$ due to the presence of noise. The relationship between blow-up in the scattering transform and noise was established for the real-valued D-bar method in,⁴³ where a bound was established on the error in the scattering transform in a disk whose radius is a function of the noise level. In the absence of noise, the scattering transform can be computed with high fidelity in a disk of large radius.^{30, 42, 43} This motivates the use of a scattering transform computed from the forward problem for the prior in an annulus outside the disk of the experimental scattering data. This approach differs significantly from the methods based on post-processing D-bar conductivity images.^{45, 46}

The original scattering data is augmented by the scattering data that corresponds to the prior outside the feasible region of computation of the true scattering data. Denoting the scattering data from the admittivity prior by S^{PR} , and the feasible region of computation by $|k| \leq R$, we form the new extended scattering data via the formula

$$S_{R,R_2}(k) := \begin{cases} S(k) & |k| \leq R \\ S^{\text{PR}}(k) & R < |k| \leq R_2 \\ 0 & R_2 < |k|, \end{cases} \quad (14)$$

where $S(k)$ is computed from current and voltage measurements using (8) for $|k| \leq R$. The truncation radius R_2 controls the amount of influence the inclusion of S^{PR} has on the reconstruction. The larger R_2 , the greater the influence. When $R_2 = R$, there is no inclusion of S^{PR} . Note that since $|S^{\text{PR}}| \rightarrow 0$ as $|k| \rightarrow \infty$, the influence of S^{PR} does not grow without bound as R_2 increases.

The second place that *a priori* information is included in the reconstruction method is in the integral forms of the D-bar equations, systems (10) and (11). The +1 and +0 terms in (10), (11) arise from terms of the form

$$\lim_{R \rightarrow \infty} \frac{1}{\pi R^2} \int_{|k| \leq R} M_{ij}(z, k) dk, i, j = 1, 2, \quad (15)$$

whose limits are 0 for M_{12} and M_{21} and 1 for M_{11} and M_{22} . Analogously to,²² to include *a priori* information encoded in the CGO solutions, the terms in (15) are replaced by a weighted integral, which we will denote by

$$M_{ij}^{\text{int}} \equiv \begin{cases} \alpha + (1 - \alpha) \int_{|k| \leq R_2} M_{ij}^{\text{PR}}(z, k) dk, i = j, \\ 0 + (1 - \alpha) \int_{|k| \leq R_2} M_{ij}^{\text{PR}}(z, k) dk, i \neq j \end{cases} \quad (16)$$

Note, when $\alpha = 1$ and $R_2 = R$ the method reduces to the original D-bar method of Section II-B without *a priori* information. Using $\alpha = 0$ corresponds to the heaviest weighting of (i.e. greatest trust in) the prior for the replacement of the asymptotic condition.

We summarize the steps of the *a priori* method. The final approximation to the admittivity is denoted by γ_{new} .

1. Setup:

Compute the DN map Λ_γ from the voltage and current measurements and determine an admittivity prior γ_{PR} .

2. Computation of Scattering Data S_{R, R_2} :

Compute the extended scattering S_{R, R_2} via (14). This involves using Step 1 of Section II-B to compute the traditional scattering data $S(k)$ for $|k| \leq R$. To obtain S^{PR} computationally, the smoothed admittivity prior is first used to compute the potential Q^{PR} via (2). Then, for $|k| \leq R_2$, the system (4) is solved, and the resulting matrix of CGO solutions is denoted by $M^{\text{PR}}(\cdot, k)$. The scattering data $S^{\text{PR}}(k)$ is then computed via (5) using Q^{PR} and M^{PR} in these equations.

3. Computation of CGO solutions:

Using the extended scattering data S_{R, R_2} , solve the systems (10) and (11) with (16) replacing the constant terms 0 and 1 to obtain CGO solutions $M_{ij}(z, k)$, $i, j = 1, 2, z \in \Omega$, where

$$\begin{aligned} M_{11}(z, k) &= M_{11}^{\text{int}} + \frac{1}{\pi k} * [M_{12}(z, \bar{k})e(z, -k)S_{21}(k)] \\ M_{12}(z, k) &= M_{12}^{\text{int}} + \frac{1}{\pi k} * [M_{11}(z, \bar{k})e(z, \bar{k})S_{12}(k)] \\ M_{21}(z, k) &= M_{21}^{\text{int}} + \frac{1}{\pi k} * [M_{22}(z, \bar{k})e(z, -k)S_{21}(k)] \\ M_{22}(z, k) &= M_{22}^{\text{int}} + \frac{1}{\pi k} * [M_{21}(z, \bar{k})e(z, \bar{k})S_{12}(k)]. \end{aligned} \quad (17)$$

4. From CGO solutions to the Admittivity $\gamma_{\text{new}}(z)$:

This is computed in the same manner as Step 3 in Section II-B to obtain γ_{new} via (12) using $M_{ij}(z, k)$, $i, j = 1, 2$, and subsequently (13).

SECTION III. Simulation and Implementation

A. Simulation of Voltage Data

The FEM was used to simulate voltages for each of the test problems using the Complete Electrode Model (CEM) on the chest-shaped domain in Figure 1 of perimeter 1016 mm, with $L = 32$ electrodes of length 22 mm and height 13.5 mm (area 297mm^2 .) The contact impedance was set to 2.4×10^{-3} on all electrodes, and trigonometric current patterns with amplitude $C = 1$ mA were used. The trigonometric current patterns are given by

$$T_\ell^j := \begin{cases} C \cos(j\theta_\ell) & 1 \leq \ell, 1 \leq j \leq \frac{L}{2} \\ C \sin((\frac{L}{2} - j)\theta_\ell) & 1 \leq \ell, \frac{L}{2} + 1 \leq j \leq L - 1, \end{cases} \quad (18)$$

where $\theta_\ell = \frac{2\pi\ell}{L}$ corresponds to the angle of the center point $z_\ell = R(\theta_\ell)e^{i\theta_\ell}$ of the ℓ -th electrode e_ℓ . The quantity T_ℓ^j therefore represents the current applied on e_ℓ corresponding to the j -th current pattern. Note that $L - 1$ linearly independent current patterns were applied since L electrodes were used in the simulations.

Zero mean Gaussian relative noise was added to each complex-valued vector of simulated voltages V^j in the same manner as ²⁵ as follows. Let η denote the desired noise level and N^j a vector of Gaussian zero mean noise that is unique for each current pattern j (and each test scenario). Then, the real and imaginary parts of the noisy voltage data $\tilde{V}^j$ were computed as

$$\begin{aligned} \Re(\tilde{V}^j) &= \Re(V^j) + \eta \max |\Re(V^j)| N^j \\ \Im(\tilde{V}^j) &= \Im(V^j) + \eta \max |\Im(V^j)| N^j. \end{aligned} \quad (19)$$

The discrete approximation Λ_γ^M to the D-N map was computed as in, ^{27, 28} where it is formed from the inner product of the vectors of current and voltage, which are discrete approximations in $\mathbb{C}^{L-1}$ and $\mathbb{C}^L$, respectively. Denoting by t_ℓ^j the (ℓ, j) -th entry of the matrix of applied currents with each column normalized with respect to the l^2 -vector norm, $t_\ell^j = \frac{T_\ell^j}{\|T^j\|_2}$, let v_ℓ^j denote the entries of the j -th voltage vector normalized so that $\sum_{\ell=1}^L v_\ell^j = 0$ and $v_\ell^j = \frac{V_\ell^j}{\|T^j\|_2}$. Let $|e_\ell|$ denote the area of the ℓ -th electrode. Then $\Lambda_\gamma^M = (R_\gamma^M)^{-1}$ where the (m, n) -th entry of R_γ^M is given by

$$R_\gamma^M(m, n) := \frac{\gamma_b}{|e_\ell|} \sum_{\ell=1}^L t_\ell^m v_\ell^n, \quad (20)$$

where γ_b denotes the background admittivity near $\partial\Omega$.

We consider two phantoms with simulated pathologies depicted in Figure 1: (b) a pneumothorax, (c) a pleural effusion. For both scenarios, the admittivity of the heart was $1.1 + 0.6i$ S/m, the lungs $0.5 + 0.4i$ S/m, and the background $0.8 + 0.4i$ S/m. The pneumothorax was set to $0.25 + 0i$ S/m and the pleural effusion to $1.1 + 0.6i$ S/m. All images are shown in DICOM orientation in which the left lung is on the viewer's right, as if viewed from the subject's feet.

B. Implementation of the a priori Method

In this paper, the admittivity prior γ_{PR} was computed using a standard D-bar reconstruction γ_{DB} recovered using Steps 1-3 of Section II-B with the measured data Λ_γ . However, in practice, any initial prior γ_{PR} can be used, making the method easily adaptable to other approaches.

1. The matrix approximation to the DN map Λ_γ was formed using the noisy voltages computed from the CEM. The admittivity prior γ_{PR} was formed as follows. First the standard D-bar reconstruction γ_{DB} was computed using Steps 1-3 of Section II-B (see ²⁵ for details regarding the computation of γ_{DB}). Next, using the "heart and lungs prior" (see the red dots of Figure 1), the maximum value of the pixels in the heart region and minimum pixel value in each lung region were computed, and the corresponding value assigned to each region to form the admittivity prior γ_{PR} . Note that the spatial prior does not assume any pathology is present. The prior γ_{PR} was then mollified to a C^1 smooth version and Q^{PR} computed using finite differences for the ∂_z and $\bar{\partial}_z$ derivatives of $\log(\gamma_{PR}(z))$.
2. The extended scattering data S_{R,R_2} was computed via (14). Using the DN maps Λ_γ and Λ_1 , the traditional scattering data $S(k)$ for $|k| \leq R$ was determined via Step 1 of Section II-B. The reader is referred to ²⁵ for the computational details of computing u_1 and u_2 and subsequently ψ_{12} and ψ_{21} . Briefly, the Fredholm integral equations for u_1 and u_1 (6) are solved by a Galerkin method, and the integrals for evaluating ψ_{12} and ψ_{21} and scattering data $S(k)$, $|k| \leq R$ in (8) are computed using a Simpson's rule, with the 32 electrode centers on the boundary as the discretization points in the quadrature. This is suitable since these are the points in z for which Ψ_{12} and Ψ_{21} are known, and since the scattering transforms are not highly oscillatory at the values of k for which they are computed here. The scattering prior S^{PR} was determined as follows. First, the admittivity prior γ_{PR} is smoothed to compute the potential Q^{PR} via (2). Then, for $|k| \leq R_2$, the system (4) was solved for M^{PR} using Fourier transforms on the following two decoupled systems:

$$\begin{cases} M_{11}^{\text{PR}}(z, k) = 1 + \frac{1}{\pi z} * [Q_{12}^{\text{PR}}(z)M_{21}^{\text{PR}}(z, k)] \\ M_{21}^{\text{PR}}(z, k) = 0 + \frac{e(z, -k)}{\pi \bar{z}} * [Q_{21}^{\text{PR}}(z)M_{11}^{\text{PR}}(z, k)] \\ M_{12}^{\text{PR}}(z, k) = 0 + \frac{e(z, \bar{k})}{\pi z} * [Q_{12}^{\text{PR}}(z)M_{22}^{\text{PR}}(z, k)] \\ M_{22}^{\text{PR}}(z, k) = 1 + \frac{1}{\pi \bar{z}} * [Q_{21}^{\text{PR}}(z)M_{12}^{\text{PR}}(z, k)], \end{cases} \quad (21)$$

where the convolutions take place in z over Ω . Using a uniform z -grid of size $2^m \times 2^m$ with stepsize h , convolutions such as $\frac{1}{\pi z} * f(z)$ can be implemented as

$$\frac{1}{\pi z} * f(z) = h^2 \text{IFFT2} \left(\text{FFT2} \left(\frac{1}{\pi z} \right) \text{FFT2}(f(z)) \right).$$

A matrix-free solution of the resulting system for each value of z was computed using GMRES with the parameters (RESTART= 10, tol= 10^{-6} , and MAXIT= 50). The system can be solved in parallel for each z in Ω . The scattering prior is then evaluated via (5) using Simpson's rule, and the combined scattering data S_{R, R_2} is subsequently formed via (14).

3. Choose a regularization weight $\alpha \in [0, 1]$. Using the combined scattering data S_{R, R_2} , the CGO solutions $M_{11}^{R_2, \alpha}$ and $M_{12}^{R_2, \alpha}$ were recovered using Fourier transforms to solve the modified equations (22)

$$\begin{cases} M_{11}^{R_2, \alpha}(z, k) = M_{11}^{\text{int}}(z) + \frac{1}{\pi k} \\ * [M_{12}^{R_2, \alpha}(z, \bar{k})e(z, -k)S_{R, R_2}^{21}(k)] \\ M_{12}^{R_2, \alpha}(z, k) = M_{12}^{\text{int}}(z) + \frac{1}{\pi k} \\ * [M_{11}^{R_2, \alpha}(z, \bar{k})e(z, \bar{k})S_{R, R_2}^{12}(k)] \end{cases} \quad (22)$$

where the convolutions take place in k over $|k| \leq R_2$ and M_{ij}^{int} computed from (16) using a Simpson's rule. An analogous system is solved to recover $M_{21}^{R_2, \alpha}$ and $M_{22}^{R_2, \alpha}$.

4. The new admittivity is recovered in the same manner as Step 3 of Section II-B to obtain γ_{new} via (12) using finite differences on $M_{ij}^{R_2, \alpha}(z, 0)$, $i, j = 1, 2$, and subsequently Fourier transforms to solve (13).

C. Examples

In this work, two noise levels were considered: 0.1% added relative noise and 1.0% relative noise. For each example, we present results with three values of the truncation radius R_2 in the prior, and three regularization weights for the D-bar equation: $\alpha = 0, 0.5, 1$. Recall that $\alpha = 0$ corresponds to the

strongest weight and $\alpha = 1$ to no weight given (see (16)). Due to the ill-posedness of the inverse problem, the radii R of admissible scattering data is problem specific, and the scattering transform will blow up in the presence of noise at a rate that is more rapid in some directions in the k -plane than others. The value chosen for each example was chosen empirically to be as large as possible without exhibiting blow up in the initial reconstruction without *a priori* information. The blow-up was more rapid in the case of 1% noise, and so in those examples a non-uniform truncation of the scattering transform was used. In such cases a threshold of the scattering data $S(k)$ was enforced by setting $S_{ij}(k) = 0$ if $|\Re(S_{ij})| > 0.15$ or $|\Im(S_{ij})| > 0.15$, where the value 0.15 was chosen empirically to be the largest permissible value of the magnitude. Determining such a threshold is intuitive from a plot of the scattering data since the blowup rate is exponential.

The admittivity prior γ_{PR} consisted of approximate knowledge of the organ boundaries (see Figure 1) with no assumption of pathology in the lungs. The values used for the prior are given in Table I.

TABLE I Admittivity Values Used in the Examples of a Simulated Pneumothorax and Pleural Effusion in the Left Lung

	Background	Left Lung	Right Lung	Heart	Pneumo.
Truth	$0.8 + 0.4i$	$0.5 + 0.2i$	$0.5 + 0.2i$	$1.1 + 0.6i$	$0.25 + 0i$
Prior 0.1%	$0.8 + 0.4i$	$0.54 + 0.16i$	$0.52 + 0.18i$	$1.04 + 0.63i$	N/A
Prior 1.0%	$0.8 + 0.4i$	$0.59 + 0.15i$	$0.55 + 0.20i$	$0.98 + 0.64i$	N/A

	Background	Left Lung	Right Lung	Heart	Pl. Effusion
Truth	$0.8 + 0.4i$	$0.5 + 0.2i$	$0.5 + 0.2i$	$1.1 + 0.6i$	$1.1 + 0.6i$
Prior 0.1%	$0.8 + 0.4i$	$0.64 + 0.25i$	$0.50 + 0.20i$	$1.08 + 0.62i$	N/A
Prior 1.0%	$0.8 + 0.4i$	$0.65 + 0.27i$	$0.46 + 0.23i$	$1.05 + 0.69i$	N/A

	Background	Left Lung	Right Lung	Heart	Pneumo.
Truth	$0.8 + 0.4i$	$0.5 + 0.2i$	$0.5 + 0.2i$	$1.1 + 0.6i$	$0.25 + 0i$
Prior 0.1%	$0.8 + 0.4i$	$0.54 + 0.16i$	$0.52 + 0.18i$	$1.04 + 0.63i$	N/A
Prior 1.0%	$0.8 + 0.4i$	$0.59 + 0.15i$	$0.55 + 0.20i$	$0.98 + 0.64i$	N/A

	Background	Left Lung	Right Lung	Heart	Pl. Effusion
Truth	$0.8 + 0.4i$	$0.5 + 0.2i$	$0.5 + 0.2i$	$1.1 + 0.6i$	$1.1 + 0.6i$
Prior 0.1%	$0.8 + 0.4i$	$0.64 + 0.25i$	$0.50 + 0.20i$	$1.08 + 0.62i$	N/A
Prior 1.0%	$0.8 + 0.4i$	$0.65 + 0.27i$	$0.46 + 0.23i$	$1.05 + 0.69i$	N/A

1) Example 1: Simulated Pneumothorax:

This test problem corresponds to phantom (b) in Figure 1. The preliminary reconstruction with no prior was computed for the 0.1% added noise case using a radius of $R = 4.5$, and for the 1% added noise case using a nonuniform truncation with a maximum radius of $R = 4.0$. Table I contains the values of the true admittivity in each region as well as the values assigned to the “heart and lungs prior” for 0.1% and 1% noise. We emphasize that we assume only approximate knowledge of the boundaries of the heart and lungs (see the red dots in Figure 1(b)), and no knowledge of the presence of a pneumothorax. Figure 2 compares the true phantom to the D-bar reconstruction with no prior, as well as to the reconstructed admittivity using the new method for the strongest use of the prior ($\alpha = 0$ and $R_2 = 9.0$) for 0.1% noise. Additional reconstructions varying the truncation radii ($R_2 = 4.5, 6.75, 9.0$) for the prior as well as weights ($\alpha = 0, 0.5, 1$) are found in Figure 3. Reconstructions for the 1% added relative noise case with truncation radii for the prior $R_2 = 4, 6, 8$ and weights $\alpha = 0, 0.5, 1$ are shown in Figure 4. Table II compares the ℓ_2 relative errors in the heart, right lung, healthy portion of the left lung, as well as the entire left lung region for each value of the parameters R_2 and α to the D-bar reconstruction with no prior. As the true

value of the pneumothorax has zero permittivity, the absolute ℓ_2 errors are given for that region rather than relative errors. Additionally, the minimum pixel value for the reconstruction in the pneumothorax region is provided.

TABLE II Errors for the Pneumothorax Example: The ℓ_2 Relative Errors in the Heart, Right Lung, Healthy Portion of the Left Lung, and Entire Left Lung are Listed. Additionally, for the Pneumothorax Region, the Absolute ℓ_2 Error is Presented as Well as the Minimum Pixel Value in the Region. Errors are Listed as (Conductivity, Permittivity)

	D-BAR RECON	$R_2 = 4.5$			$R_2 = 6.75$			$R_2 = 9$		
		$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$
0.1% Noise										
HEART	(22%, 29%)	(23%, 30%)	(18%, 22%)	(14%, 23%)	(22%, 28%)	(19%, 23%)	(16%, 18%)	(21%, 25%)	(19%, 21%)	(17%, 18%)
R. LUNG	(29%, 52%)	(29%, 51%)	(19%, 34%)	(15%, 29%)	(26%, 43%)	(20%, 32%)	(16%, 27%)	(22%, 36%)	(19%, 31%)	(17%, 29%)
L. LUNG (HEALTHY)	(45%, 67%)	(44%, 66%)	(32%, 45%)	(23%, 37%)	(36%, 53%)	(29%, 42%)	(23%, 36%)	(34%, 52%)	(30%, 45%)	(26%, 40%)
L. LUNG (ENTIRE)	(59%, 91%)	(58%, 90%)	(46%, 69%)	(37%, 58%)	(49%, 73%)	(42%, 63%)	(36%, 57%)	(47%, 72%)	(42%, 66%)	(39%, 62%)
PNEUMO (ABS ℓ_2)	(3.02, 1.85)	(3.00, 1.83)	(2.61, 1.56)	(2.31, 1.35)	(2.66, 1.54)	(2.41, 1.41)	(2.21, 1.31)	(2.54, 1.50)	(2.38, 1.44)	(2.25, 1.40)
PNEUMO (MIN)	(0.61, 0.19)	(0.61, 0.19)	(0.51, 0.12)	(0.44, 0.07)	(0.54, 0.13)	(0.47, 0.09)	(0.42, 0.06)	(0.49, 0.11)	(0.45, 0.08)	(0.41, 0.05)

	D-BAR	$R_2 = 4$			$R_2 = 6$			$R_2 = 8$		
	RECON	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$
0.1% Noise										
HEART	(25%, 30%)	(25%, 30%)	(20%, 23%)	(17%, 28%)	(24%, 27%)	(21%, 23%)	(19%, 20%)	(23%, 24%)	(22%, 19%)	(20%, 16%)
R. LUNG	(29%, 48%)	(28%, 48%)	(18%, 33%)	(13%, 30%)	(23%, 39%)	(17%, 29%)	(13%, 25%)	(21%, 33%)	(17%, 28%)	(15%, 25%)
L. LUNG (HEALTHY)	(43%, 54%)	(42%, 53%)	(32%, 40%)	(24%, 48%)	(36%, 42%)	(30%, 39%)	(25%, 44%)	(33%, 41%)	(29%, 40%)	(26%, 41%)
L. LUNG (ENTIRE)	(59%, 88%)	(59%, 86%)	(49%, 67%)	(41%, 65%)	(53%, 69%)	(46%, 61%)	(41%, 61%)	(50%, 67%)	(45%, 63%)	(42%, 62%)
PNEUMO (ABS ℓ_2)	(3.25, 2.11)	(3.31, 2.05)	(2.93, 1.65)	(2.63, 1.35)	(3.02, 1.66)	(2.75, 1.43)	(2.53, 1.26)	(2.93, 1.60)	(2.74, 1.47)	(2.58, 1.37)
PNEUMO (MIN)	(0.66, 0.20)	(0.67, 0.21)	(0.59, 0.14)	(0.51, 0.07)	(0.62, 0.16)	(0.55, 0.10)	(0.49, 0.05)	(0.58, 0.12)	(0.53, 0.08)	(0.49, 0.05)

	D-BAR	$R_2 = 4.5$			$R_2 = 6.75$			$R_2 = 9$		
	RECON	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$
0.1% Noise										
HEART	(22%, 29%)	(23%, 30%)	(18%, 22%)	(14%, 23%)	(22%, 28%)	(19%, 23%)	(16%, 18%)	(21%, 25%)	(19%, 21%)	(17%, 18%)
R. LUNG	(29%, 52%)	(29%, 51%)	(19%, 34%)	(15%, 29%)	(26%, 43%)	(20%, 32%)	(16%, 27%)	(22%, 36%)	(19%, 31%)	(17%, 29%)
L. LUNG (HEALTHY)	(45%, 67%)	(44%, 66%)	(32%, 45%)	(23%, 37%)	(36%, 53%)	(29%, 42%)	(23%, 36%)	(34%, 52%)	(30%, 45%)	(26%, 40%)
L. LUNG (ENTIRE)	(59%, 91%)	(58%, 90%)	(46%, 69%)	(37%, 58%)	(49%, 73%)	(42%, 63%)	(36%, 57%)	(47%, 72%)	(42%, 66%)	(39%, 62%)
PNEUMO (ABS ℓ_2)	(3.02, 1.85)	(3.00, 1.83)	(2.61, 1.56)	(2.31, 1.35)	(2.66, 1.54)	(2.41, 1.41)	(2.21, 1.31)	(2.54, 1.50)	(2.38, 1.44)	(2.25, 1.40)
PNEUMO (MIN)	(0.61, 0.19)	(0.61, 0.19)	(0.51, 0.12)	(0.44, 0.07)	(0.54, 0.13)	(0.47, 0.09)	(0.42, 0.06)	(0.49, 0.11)	(0.45, 0.08)	(0.41, 0.05)

	D-BAR	$R_2 = 4$			$R_2 = 6$			$R_2 = 8$		
	RECON	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$
0.1% Noise										
HEART	(25%, 30%)	(25%, 30%)	(20%, 23%)	(17%, 28%)	(24%, 27%)	(21%, 23%)	(19%, 20%)	(23%, 24%)	(22%, 19%)	(20%, 16%)
R. LUNG	(29%, 48%)	(28%, 48%)	(18%, 33%)	(13%, 30%)	(23%, 39%)	(17%, 29%)	(13%, 25%)	(21%, 33%)	(17%, 28%)	(15%, 25%)
L. LUNG (HEALTHY)	(43%, 54%)	(42%, 53%)	(32%, 40%)	(24%, 48%)	(36%, 42%)	(30%, 39%)	(25%, 44%)	(33%, 41%)	(29%, 40%)	(26%, 41%)
L. LUNG (ENTIRE)	(59%, 88%)	(59%, 86%)	(49%, 67%)	(41%, 65%)	(53%, 69%)	(46%, 61%)	(41%, 61%)	(50%, 67%)	(45%, 63%)	(42%, 62%)
PNEUMO (ABS ℓ_2)	(3.25, 2.11)	(3.31, 2.05)	(2.93, 1.65)	(2.63, 1.35)	(3.02, 1.66)	(2.75, 1.43)	(2.53, 1.26)	(2.93, 1.60)	(2.74, 1.47)	(2.58, 1.37)
PNEUMO (MIN)	(0.66, 0.20)	(0.67, 0.21)	(0.59, 0.14)	(0.51, 0.07)	(0.62, 0.16)	(0.55, 0.10)	(0.49, 0.05)	(0.58, 0.12)	(0.53, 0.08)	(0.49, 0.05)

Fig. 2. Reconstructions for the pneumothorax example with 0.1% noise plotted on the same scale. Left: the true admittivity. Middle: the initial D-bar reconstruction γ_{DB} . Right: the new admittivity γ_{new} with strongest prior $R_2 = 9.0$ and $\alpha = 0$ all plotted on the same scale.

Conductivity

Permittivity

Fig. 3. Reconstructions of simulated pneumothorax with 0.1% added noise. Regularization parameter $\alpha = 1, 0.5, 0$ increases the influence of M_{ij}^{int} as α decreases, and $R_2 = 4.5, 6.75, 9.0$ (rows) increases the influence of S^{PR} as R_2 increases. No pneumothorax is assumed to be present in the prior. Max conductivity: 1.1608, max permittivity: 0.7983.

Conductivity

Permittivity

Fig. 4. Reconstructions of simulated pneumothorax with 1.0% added noise. Regularization parameter $\alpha = 1, 0.5, 0$ increases the influence of M_{ij}^{int} as α decreases, and $R_2 = 4, 6, 8$ (rows) increases the influence of S^{PR} as R_2 increases. No pneumothorax is assumed to be present in the prior. Max conductivity: 1.1187, max permittivity: 0.8820.

2) Example 2: Simulated Pleural Effusion:

This test problem corresponds to phantom (c) in Figure 1. The preliminary reconstruction with no prior was computed for the 0.1% added noise case using a radius of $R = 5.5$, and for the 1% added noise case using

a nonuniform truncation with a maximum radius of $R = 4.5$. Table I presents the values used in the prior γ_{PR} for each noise level and Figure 5 compares the true phantom to reconstructions using the D-bar method with no prior and the new method for the strongest weight of the prior at 0.1% added noise. Reconstructions for additional truncation radii ($R_2 = 5.5, 8.25, 11$) and weights ($\alpha = 0, 0.5, 1$) of the prior are found in Figure 6. Reconstructions for the 1% added relative noise case with truncation radii for the prior $R_2 = 4.5, 6.75, 9.0$ and weights $\alpha = 0, 0.5, 1$ are shown in Figure 7. Table III compares the ℓ_2 relative errors of the D-bar reconstruction with no prior to those of the new method in each region, and also gives the maximum value in the pleural effusion region.

TABLE III Errors for the Pleural Effusion Example: The ℓ_2 Relative Errors in Each Region as Well as the Maximum Pixel Value in the Region Containing the Pleural Effusion. Errors are Listed as (Conductivity, Permittivity)

	D-bar Recon	$R_2 = 5.5$			$R_2 = 8.25$			$R_2 = 11$		
		$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$
0.1% Noise										
HEART	(20%, 28%)	(21%, 27%)	(17%, 22%)	(13%, 19%)	(20%, 24%)	(18%, 21%)	(15%, 18%)	(20%, 24%)	(18%, 22%)	(17%, 20%)
R. LUNG	(26%, 51%)	(25%, 50%)	(18%, 37%)	(15%, 29%)	(20%, 41%)	(17%, 35%)	(15%, 31%)	(19%, 39%)	(17%, 36%)	(16%, 34%)
L. LUNG (HEALTHY)	(44%, 71%)	(43%, 69%)	(36%, 55%)	(29%, 44%)	(39%, 64%)	(35%, 58%)	(31%, 52%)	(37%, 62%)	(34%, 58%)	(32%, 54%)
L. LUNG (ENTIRE)	(30%, 34%)	(30%, 34%)	(31%, 34%)	(32%, 37%)	(30%, 33%)	(31%, 34%)	(31%, 35%)	(30%, 33%)	(30%, 33%)	(30%, 34%)
PL. EFFUSION	(25%, 25%)	(26%, 24%)	(29%, 30%)	(33%, 36%)	(28%, 26%)	(30%, 29%)	(31%, 32%)	(28%, 26%)	(29%, 28%)	(30%, 30%)
PL. EFFUSION (MAX)	(0.92, 0.52)	(0.91, 0.53)	(0.87, 0.49)	(0.83, 0.45)	(0.91, 0.54)	(0.88, 0.51)	(0.85, 0.50)	(0.87, 0.53)	(0.86, 0.53)	(0.84, 0.54)
1.0% Noise										
	D-bar	$R_2 = 4.5$			$R_2 = 6.75$			$R_2 = 9$		
HEART	(20%, 22%)	(19%, 20%)	(15%, 22%)	(13%, 34%)	(19%, 21%)	(16%, 20%)	(14%, 22%)	(18%, 17%)	(16%, 17%)	(14%, 20%)
R. LUNG	(23%, 60%)	(22%, 60%)	(16%, 45%)	(19%, 37%)	(18%, 53%)	(15%, 44%)	(15%, 38%)	(14%, 48%)	(14%, 43%)	(15%, 39%)
L. LUNG (HEALTHY)	(41%, 76%)	(39%, 73%)	(31%, 56%)	(23%, 44%)	(35%, 67%)	(29%, 58%)	(25%, 51%)	(32%, 65%)	(29%, 60%)	(26%, 56%)
L. LUNG (ENTIRE)	(32%, 37%)	(33%, 38%)	(34%, 40%)	(37%, 44%)	(33%, 38%)	(34%, 39%)	(35%, 41%)	(33%, 38%)	(33%, 39%)	(34%, 40%)
PL. EFFUSION	(29%, 26%)	(31%, 30%)	(35%, 37%)	(40%, 44%)	(33%, 32%)	(35%, 36%)	(38%, 40%)	(33%, 32%)	(35%, 35%)	(36%, 37%)
PL. EFFUSION (MAX)	(0.83, 0.48)	(0.84, 0.45)	(0.80, 0.40)	(0.76, 0.37)	(0.86, 0.46)	(0.83, 0.43)	(0.79, 0.40)	(0.84, 0.46)	(0.81, 0.44)	(0.79, 0.44)
0.1% Noise										
	D-bar Recon	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$	$\alpha = 1$	$\alpha = 0.5$	$\alpha = 0$
HEART	(20%, 28%)	(21%, 27%)	(17%, 22%)	(13%, 19%)	(20%, 24%)	(18%, 21%)	(15%, 18%)	(20%, 24%)	(18%, 22%)	(17%, 20%)
R. LUNG	(26%, 51%)	(25%, 50%)	(18%, 37%)	(15%, 29%)	(20%, 41%)	(17%, 35%)	(15%, 31%)	(19%, 39%)	(17%, 36%)	(16%, 34%)
L. LUNG (HEALTHY)	(44%, 71%)	(43%, 69%)	(36%, 55%)	(29%, 44%)	(39%, 64%)	(35%, 58%)	(31%, 52%)	(37%, 62%)	(34%, 58%)	(32%, 54%)
L. LUNG (ENTIRE)	(30%, 34%)	(30%, 34%)	(31%, 34%)	(32%, 37%)	(30%, 33%)	(31%, 34%)	(31%, 35%)	(30%, 33%)	(30%, 33%)	(30%, 34%)
PL. EFFUSION	(25%, 25%)	(26%, 24%)	(29%, 30%)	(33%, 36%)	(28%, 26%)	(30%, 29%)	(31%, 32%)	(28%, 26%)	(29%, 28%)	(30%, 30%)
PL. EFFUSION (MAX)	(0.92, 0.52)	(0.91, 0.53)	(0.87, 0.49)	(0.83, 0.45)	(0.91, 0.54)	(0.88, 0.51)	(0.85, 0.50)	(0.87, 0.53)	(0.86, 0.53)	(0.84, 0.54)
1.0% Noise										
	D-bar	$R_2 = 4.5$			$R_2 = 6.75$			$R_2 = 9$		
HEART	(20%, 22%)	(19%, 20%)	(15%, 22%)	(13%, 34%)	(19%, 21%)	(16%, 20%)	(14%, 22%)	(18%, 17%)	(16%, 17%)	(14%, 20%)
R. LUNG	(23%, 60%)	(22%, 60%)	(16%, 45%)	(19%, 37%)	(18%, 53%)	(15%, 44%)	(15%, 38%)	(14%, 48%)	(14%, 43%)	(15%, 39%)
L. LUNG (HEALTHY)	(41%, 76%)	(39%, 73%)	(31%, 56%)	(23%, 44%)	(35%, 67%)	(29%, 58%)	(25%, 51%)	(32%, 65%)	(29%, 60%)	(26%, 56%)
L. LUNG (ENTIRE)	(32%, 37%)	(33%, 38%)	(34%, 40%)	(37%, 44%)	(33%, 38%)	(34%, 39%)	(35%, 41%)	(33%, 38%)	(33%, 39%)	(34%, 40%)
PL. EFFUSION	(29%, 26%)	(31%, 30%)	(35%, 37%)	(40%, 44%)	(33%, 32%)	(35%, 36%)	(38%, 40%)	(33%, 32%)	(35%, 35%)	(36%, 37%)
PL. EFFUSION (MAX)	(0.83, 0.48)	(0.84, 0.45)	(0.80, 0.40)	(0.76, 0.37)	(0.86, 0.46)	(0.83, 0.43)	(0.79, 0.40)	(0.84, 0.46)	(0.81, 0.44)	(0.79, 0.44)

Fig. 5. Reconstructions for the pleural effusion example with 0.1% noise plotted on the same scale. Left: the true admittance. Middle: the initial D-bar reconstruction γ_{DB} . Right: the new admittance γ_{new} with strongest prior $R_2 = 11.0$ and $\alpha = 0$ all plotted on the same scale maxed out at $1.1 + 0.6i$.

Conductivity

Permittivity

Fig. 6. Reconstructions of simulated pleural effusion with 0.1% added noise. Regularization parameter $\alpha = 1, 0.5, 0$ increases the influence of M_{ij}^{int} as α decreases, and $R_2 = 5.5, 8.25, 11$ (rows) increases the influence of S^{PR} as R_2 increases. No effusion is assumed to be present in the prior. Max conductivity: 1.1537, max permittivity 0.7416.

Conductivity

Permittivity

Fig. 7. Reconstructions of simulated pleural effusion with 1.0% added noise. Regularization parameter $\alpha = 1, 0.5, 0$ increases the influence of M_{ij}^{int} as α decreases, and $R_2 = 4.5, 6.75, 9.0$ (rows) increases the influence of S^{PR} as R_2 increases. No effusion is assumed to be present in the prior. Max conductivity: 1.2452, max permittivity: 0.9699.

SECTION IV. Discussion and Conclusions

In Figures 3, 4, 6, and 7, the upper left figure corresponds to the weakest/no prior, and the weight of the prior increases to the right and down the columns. It is evident that the spatial resolution of the organ boundaries improves with the introduction of the prior and as the influence of the prior increases. The colormaps are maxed out at 1.1 for the conductivity and 0.6 for the permittivity to allow easier comparative viewing. The maximum conductivity and permittivity values always occurred for the smallest R_2 and strongest weight $\alpha = 0$. In the case of the pneumothorax, no pathology is evident without the inclusion of a prior, but as the influence of the prior increases, even though the prior includes no assumption of pathology, the pneumothorax is clearly visible in the reconstructions. However, in both the conductivity and permittivity images, a lower conductivity and permittivity region becomes evident in the dorsal right lung as well, which is an artifact of the reconstruction, and it becomes stronger as the weighting of M_{ij}^{int} increases ($\alpha = 0.5$ and 0). This artifact is less pronounced in the permittivity images, and is arguably not present in the 1.0% added noise case in the permittivity images. The error norms in Table II clearly demonstrate that the introduction of the prior decreases the ℓ_2 errors for all regions of interest, as well as the minimum value in the pneumothorax region.

The presence of the simulated pleural effusion, on the other hand, is clearly evident in the reconstructions with the weakest/no prior for both conductivity and permittivity and for both noise levels. The presence of the prior improves the spatial resolution of the organs and the region of the effusion, but since the regularization results in reconstructed conductivity and permittivity functions that are smooth, there is a smooth transition from the healthy ventral portion of the left lung to the effusion, and so the boundary is far from being as sharp as in the piecewise constant phantom. In practice, image segmentation is often used on reconstructed EIT images, which would likely improve the appearance of the reconstructed images. Alternatively, once a pathology is visible, an iterative method could then be invoked as in ²² which segments the prior in the region of a possible pathology potentially sharpening the pathology even more. Post-processing approaches are left for future work. The error norms in Table III show that employing the prior greatly decreases the ℓ_2 errors for the healthy portions of the phantom and do not have a large effect on the pleural effusion region even though the prior assumes the entire left lung has an admittivity (lower rather than higher) than the background.

Although the error norms are, on average, higher for reconstructions from the higher noise levels, they are reduced by the introduction of the prior. A stronger weighting may be required to achieve an equivalent accuracy to the lower noise level. For each given α and R_2 , the algorithm took approximately 35 seconds to recover the updated admittivity γ_{new} . The implementation was not optimized for speed and thus further speedup is attainable.

It is clear from all of these images and error norms that this method is highly effective when organ boundaries are known with some confidence for improving the reconstructions without any bias of prior knowledge of the pathology. The influence of various qualities of prior knowledge of the boundary and organ boundaries is left for future work, as are results from experimental data. In practice, this high quality knowledge of organ boundaries corresponds to electrodes placed in the same plane as a CT scan slice. This

can be accomplished with careful use of fiducial markers, and averaging of several slices to account for the fact that EIT electrodes are typically much higher than a CT scan slice, resulting in an image that corresponds to a much thicker slice.

References

1. G. Alessandrini, M. Di Cristo, "Stable determination of an inclusion by boundary measurements", *SIAM J. Math. Anal.*, vol. 37, no. 1, pp. 200-217, 2005.
2. G. Alessandrini, E. Cabib, "EIT and the average conductivity", *J. Inverse Ill-Posed Problems*, vol. 16, no. 8, pp. 727-736, Jan. 2008.
3. J. A. Barceló, T. Barceló, A. Ruiz, "Stability of the inverse conductivity problem in the plane for less regular conductivities", *J. Differ. Equ.*, vol. 173, no. 2, pp. 231-270, 2001.
4. T. Barcelo, D. Faraco, A. Ruiz, "Stability of Calderón inverse conductivity problem in the plane", *J. Math. Pures Appl.*, vol. 88, no. 6, pp. 522-556, 2007.
5. C. Grant, T. Pham, J. Hough, T. Riedel, C. Stocker, A. Schibler, "Measurement of ventilation and cardiac related impedance changes with electrical impedance tomography", *Critical Care*, vol. 15, pp. R37, Jan. 2011.
6. A. Schlibler, T. Pham, A. Moray, C. Stocker, "Ventilation and cardiac related impedance changes in children undergoing corrective open heart surgery", *Physiol. Meas.*, vol. 34, pp. 1319-1327, 2013.
7. H. Reinius et al., "Real-time ventilation and perfusion distributions by electrical impedance tomography during one-lung ventilation with capnothorax", *Acta Anaesthesiol. Scand.*, vol. 59, no. 3, pp. 354-368, 2015.
8. C. Karagiannidis et al., "Position-dependent distribution of ventilation measured with electrical impedance tomography", *Eur. Respirat. J.*, vol. 46, no. 59, pp. PA2144, Sep. 2015.
9. G. Cinnella et al., "Physiological effects of the open lung approach in patients with early mild diffuse acute respiratory distress syndrome: an electrical impedance tomography study", *Anesthesiologists*, vol. 123, no. 5, pp. 1113-1121, 2015.
10. A. Pesenti et al., "Imaging in acute respiratory distress syndrome", *Intensive Care Med.*, vol. 42, no. 5, pp. 686-698, 2016.
11. E. L. V. Costa et al., "Bedside estimation of recruitable alveolar collapse and hyperdistension by electrical impedance tomography", *Intensive Care Med.*, vol. 35, no. 6, pp. 1132-1137, 2009.
12. N. J. Avis, D. C. Barber, "Incorporating a priori information into the Sheffield filtered backprojection algorithm", *Physiol. Meas.*, vol. 16, no. 3A, pp. A111-A122, 1995.
13. U. Baysal, B. Eyüboğlu, "Use of a priori information in estimating tissue resistivities—A simulation study", *Phys. Med. Biol.*, vol. 43, no. 12, pp. 3589-3606, 1998.
14. E. Camargo, "Development of an absolute electrical impedance imaging algorithm for clinical use", 2013.
15. H. Dehghani, D. Barber, I. Basarab-Horwath, "Incorporating a priori anatomical information into image reconstruction in electrical impedance tomography", *Physiol. Meas.*, vol. 20, no. 1, pp. 87-102, 1999.
16. D. C. Dobson, F. Santosa, "An image-enhancement technique for electrical impedance tomography", *Inverse Problems*, vol. 10, no. 2, pp. 317-334, 1994.
17. D. Ferrario, B. Grychtol, A. Adler, J. Sola, S. H. Böhm, M. Bodenstein, "Toward morphological thoracic EIT: Major signal sources correspond to respective organ locations in CT", *IEEE Trans. Biomed. Eng.*, vol. 59, no. 11, pp. 3000-3008, Nov. 2012.
18. D. Flores-Tapia, S. Pistorius, "Electrical impedance tomography reconstruction using a monotonicity approach based on a priori knowledge", *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc. (EMBC)*, pp. 4996-4999, Aug. 2010.

19. J. Kaipio, V. Kolehmainen, M. Vauhkonen, E. Somersalo, "Inverse problems with structural prior information", *Inverse Problems*, vol. 15, no. 3, pp. 713-729, 1999.
20. M. Soleimani, "Electrical impedance tomography imaging using a priori ultrasound data", *Biomed. Eng. OnLine*, vol. 5, no. 8, Feb. 2006.
21. M. Vauhkonen, D. Vadasz, P. A. Karjalainen, E. Somersalo, J. P. Kaipio, "Tikhonov regularization and prior information in electrical impedance tomography", *IEEE Trans. Med. Imag.*, vol. 17, no. 2, pp. 285-293, Apr. 1998.
22. M. Alsaker, J. L. Mueller, "A D-bar algorithm with a priori information for 2-D electrical impedance tomography", *SIAM J. Imag. Sci.*, 2016.
23. S. J. Hamilton, "A direct D-bar reconstruction algorithm for complex admittivities in $\mathbb{R}^2$ for the 2-D EIT problem", 2012.
24. S. Hamilton, C. Herrera, J. L. Mueller, A. VonHerrmann, "A direct D-bar reconstruction algorithm for recovering a complex conductivity in 2-D", *Inverse Problems*, vol. 28, no. 9, pp. 095005, 2012.
25. S. J. Hamilton, J. L. Mueller, "Direct EIT reconstructions of complex admittivities on a chest-shaped domain in 2-D", *IEEE Trans. Med. Imag.*, vol. 32, no. 4, pp. 757-769, Apr. 2013.
26. M. Dodd, J. L. Mueller, "A real-time d-bar algorithm for 2-D electrical impedance tomography data", *Inverse Problems Imag.*, vol. 8, no. 4, pp. 1013-1031, 2014.
27. D. Isaacson, J. L. Mueller, J. C. Newell, S. Siltanen, "Reconstructions of chest phantoms by the D-bar method for electrical impedance tomography", *IEEE Trans. Med. Imag.*, vol. 23, no. 7, pp. 821-828, Jul. 2004.
28. D. Isaacson, J. L. Mueller, J. C. Newell, S. Siltanen, "Imaging cardiac activity by the D-bar method for electrical impedance tomography", *Physiol. Meas.*, vol. 27, no. 5, pp. S43-S50, 2006.
29. E. K. Murphy, J. L. Mueller, "Effect of domain-shape modeling and measurement errors on the 2-D D-bar method for electrical impedance tomography", *IEEE Trans. Med. Imag.*, vol. 28, no. 10, pp. 1576-1584, Oct. 2009.
30. S. Siltanen, J. Mueller, D. Isaacson, "An implementation of the reconstruction algorithm of a Nachman for the 2-D inverse conductivity problem", *Inverse Problems*, vol. 16, no. 3, pp. 681-699, 2000.
31. A. I. Nachman, "Global uniqueness for a two-dimensional inverse boundary value problem", *Ann. Math.*, vol. 143, no. 1, pp. 71-96, 1996.
32. F. C. Trigo, R. Gonzalez-Lima, M. B. P. Amato, "Electrical impedance tomography using the extended Kalman filter", *IEEE Trans. Biomed. Eng.*, vol. 51, no. 1, pp. 72-81, Jan. 2004.
33. V. Kolehmainen, M. Lassas, P. Ola, S. Siltanen, "Recovering boundary shape and conductivity in electrical impedance tomography", *Inverse Problems Imag.*, vol. 7, no. 1, pp. 217-242, 2013.
34. C. Herrera, M. Vallejo, J. Mueller, R. Lima, "Direct 2-D reconstructions of conductivity and permittivity from EIT data on a human chest", *IEEE Trans. Med. Imag.*, vol. 34, no. 1, pp. 267-274, Jan. 2015.
35. K.-S. Cheng, D. Isaacson, J. C. Newell, D. G. Gisser, "Electrode models for electric current computed tomography", *IEEE Trans. Biomed. Eng.*, vol. 36, no. 9, pp. 918-924, Sep. 1989.
36. E. Somersalo, M. Cheney, D. Isaacson, "Existence and uniqueness for electrode models for electric current computed tomography", *SIAM J. Appl. Math.*, vol. 52, no. 4, pp. 1023-1040, 1992.
37. E. Francini, "Recovering a complex coefficient in a planar domain from Dirichlet-to-Neumann map", *Inverse Problems*, vol. 16, no. 1, pp. 107-119, 2000.
38. R. M. Brown, G. Uhlmann, "Uniqueness in the inverse conductivity problem for nonsmooth conductivities in two dimensions", *Commun. Partial Differ. Equ.*, vol. 22, no. 5, pp. 1009-1027, 1997.
39. K. Knudsen, A. Tamasan, "Reconstruction of less regular conductivities in the plane", *Commun. Partial Differ. Equ.*, vol. 29, no. 3, pp. 361-381, 2004.

- 40. K. Knudsen, "A new direct method for reconstructing isotropic conductivities in the plane", *Physiol. Meas.*, vol. 24, no. 2, pp. 391-403, 2003.
- 41. L. D. Faddeev, "Increasing solutions of the Schrödinger equation", *Soviet Phys. Doklady*, vol. 10, pp. 1033-1035, 1966.
- 42. J. Mueller, S. Siltanen, *Linear Nonlinear Inverse Problems With Practical Application*, Philadelphia, PA, USA:SIAM, 2012.
- 43. K. Knudsen, M. Lassas, J. Mueller, S. Siltanen, "Regularized D-bar method for the inverse conductivity problem", *Inverse Problems Imag.*, vol. 3, no. 4, pp. 599-624, 2009.
- 44. K. Knudsen, J. Mueller, S. Siltanen, "Numerical solution method for the Dbar-equation in the plane", *J. Comput. Phys.*, vol. 198, no. 2, pp. 500-517, 2004.
- 45. S. J. Hamilton, A. Hauptmann, S. Siltanen, "A data-driven edge-preserving D-bar method for electrical impedance tomography", *Inverse Problems Imag.*, vol. 8, no. 4, pp. 1053-1072, 2014.
- 46. S. J. Hamilton, J. M. Reyes, S. Siltanen, X. Zhang, "A hybrid segmentation and D-bar method for Electrical Impedance Tomography", *SIAM J. Imag. Sci.*, vol. 9, no. 2, pp. 770-793, 2016.