Marquette University

e-Publications@Marquette

Biomedical Engineering Faculty Research and Publications

Biomedical Engineering, Department of

2017

Smartphone Based Optical Spectrometer for Diffusive Reflectance Spectroscopic Measurement of Hemoglobin

Perry Edwards *Atoptix, Inc.*

Chenji Zhang Pennsylvania State University

Baigang Zhang Atoptix, Inc.

Xiangqian Hong Marquette University

Vivek K. Nagarajan The University of Akron

See next page for additional authors

Follow this and additional works at: https://epublications.marquette.edu/bioengin_fac

Part of the Biomedical Engineering and Bioengineering Commons

Recommended Citation

Edwards, Perry; Zhang, Chenji; Zhang, Baigang; Hong, Xiangqian; Nagarajan, Vivek K.; Yu, Bing; and Liu, Zhiwen, "Smartphone Based Optical Spectrometer for Diffusive Reflectance Spectroscopic Measurement of Hemoglobin" (2017). *Biomedical Engineering Faculty Research and Publications*. 641. https://epublications.marquette.edu/bioengin_fac/641

Authors

Perry Edwards, Chenji Zhang, Baigang Zhang, Xiangqian Hong, Vivek K. Nagarajan, Bing Yu, and Zhiwen Liu

SCIENTIFIC REPORTS

Received: 4 July 2017 Accepted: 7 September 2017 Published online: 22 September 2017

OPEN Smartphone based optical spectrometer for diffusive reflectance spectroscopic measurement of hemoglobin

Perry Edwards¹, Chenji Zhang², Baigang Zhang¹, Xiangqian Hong³, Vivek K. Nagarajan⁴, Bing Yu³ & Zhiwen Liu²

We report a miniature, visible to near infrared G-Fresnel spectrometer that contains a complete spectrograph system, including the detection hardware and connects with a smartphone through a microUSB port for operational control. The smartphone spectrometer is able to achieve a resolution of ~5 nm in a wavelength range from 400 nm to 1000 nm. We further developed a diffuse reflectance spectroscopy system using the smartphone spectrometer and demonstrated the capability of hemoglobin measurement. Proof of concept studies of tissue phantoms yielded a mean error of 9.2% on hemoglobin concentration measurement, comparable to that obtained with a commercial benchtop spectrometer. The smartphone G-Fresnel spectrometer and the diffuse reflectance spectroscopy system can potentially enable new point-of-care opportunities, such as cancer screening.

Hemoglobin is an important biomarker for diagnosing a variety of clinical conditions¹⁻⁸. For instance, angiogenesis and hypoxia are two salient features of tumor growth. Thus, non-invasive measurement of both oxy- and deoxy-hemoglobin in tissue, from which the total hemoglobin concentration and its oxygenation state can be determined, is a promising approach for early detection and screening of breast cancer⁹⁻¹¹, oral cancer¹²⁻¹⁵, cervical cancer¹⁶⁻¹⁸, and many other malignancies. Therefore, there is an important need for developing a portable and cost-efficient tool to perform non-invasive measurement of hemoglobin at local tissues for point of care (POC) applications. Note that current gold standard for diagnosing these conditions is biopsy, which is invasive, costly, and not easily accessible in resource scarce communities and developing regions.

Diffuse reflectance spectroscopy (DRS) is a well-established method to measure hemoglobin in tissue ^{12,13,19-23} Unlike specular reflection, in DRS light penetrates deep into a tissue and re-emerges to the surface only after undergoing interactions with the tissue, such as absorption by chromophores and multiple scatterings by cellular and intercellular structures. As such, the diffuse reflectance spectrum contains important information of chromophore concentrations and tissue scattering properties. In the visible wavelength, the most significant absorbers in human tissue are oxy-hemoglobin and deoxy-hemoglobin. Due to its non-invasive nature, DRS can avoid blood drawing and consumables, and thus lends itself to significant cost reduction during usage. Yet, conventional implementation of DRS typically involves the use of bulky and costly optical spectrometers, which remain a significant hindrance for gaining wider accessibility, especially in resource-limited areas.

In recent years, there is an increasing interest in developing portable, low-cost optical spectroscopic devices and integrating them with mobile phones to leverage the ease of accessing and processing data using mobile computing technology.^{24–29} For instance, several smartphone based spectrometers were demonstrated with wave-length resolutions ranging from 2 to 15 nm^{25,28,30,31}. In large part due to the competing requirements of size, cost, and performance, it remains challenging to realize a truly miniature, cost-effective, and high-performance mobile phone spectrometer. Developing these capabilities is critical for enabling POC applications that demand high accuracy in measurements at low cost. In the past several years, G-Fresnel, a diffractive optical element with the

¹Atoptix, Inc., 200 Innovation Blvd., Suite, 234–1, 16803, State College, PA, USA. ²Department of Electrical Engineering and Materials Research Institute, The Pennsylvania State University, University Park, 16802, PA, USA. ³Department of Biomedical Engineering, Marguette University and Medical College of Wisconsin, Milwaukee, WI, 53233, United States. ⁴Department of Biomedical Engineering, The University of Akron, Akron, OH, 44325, United States. Correspondence and requests for materials should be addressed to Z.L. (email: zliu@psu.edu)



Figure 1. (a) A schematic of the G-Fresnel spectrometer; (b) a G-Fresnel spectrometer attached to a smartphone displaying the measured spectrum of a calibration lamp; and (c) measured spectrum of a tunable narrow-band calibration source showing the wavelength resolution at several representative wavelengths.

dual functionality of focusing and dispersion has been developed for implementing miniaturized high-resolution optical spectrometers³²⁻³⁵. A transmission G-Fresnel, as used in this work, can be fabricated via soft lithography, by sandwiching polydimethylsiloxane (PDMS) prepolymer between a negative Fresnel lens mold and a grating mold³⁵. As a result, the fabricated device consists of a Fresnel lens pattern in one side to focus or collimate impinging light, and a grating pattern on the other side to disperse the different constituent wavelengths. A smartphone spectrometer was recently demonstrated, by directly positioning a G-Fresnel in front of the built-in camera of a smartphone; application to Bradford assay of protein concentration was also studied³⁶. Albeit ultralow cost, using the built-in camera as a detector, however, both poses alignment challenges to accommodate the ever-changing positions of the back-camera across different smartphone models, and necessitates workarounds to undo the built-in image processing that often wrecks havoc on spectral images aquired using a smartphone camera. To overcome the challenge, in this work, we developed a G-Fresnel spectrometer that contains a complete spectrograph system, including the detection hardware and control electronics, and can connect with a smartphone through a microUSB port as an add-on device for operational control. We further developed a DRS system based on this smartphone spectrometer and demonstrated quantitive measurement of hemoglobin concentration and reduced scattering coefficient through tissue phantom study. The smartphone spectrometer based DRS system has the potential to be broadly implemented in resource-limited regions.

Materials and Methods

Spectrometer design and fabrication. A schematic diagram of our G-Fresnel spectrometer is shown in Fig. 1a. An incoming fibre delivers light signal onto a slit. A transmission G-Fresnel device, comprising a grating pattern in one side (600 lines per mm) and Fresnel lens pattern on the other side (focal length: 1 inch at 588 nm), both collimates the light emanating from the slit and disperses the different wavelengths across a mobile camera system containing a built-in lens that focuses the dispersed light onto a complementary metal oxide semiconductor (CMOS) image sensor (Omnivision). The fabrication of G-Fresnel was described in details elsewhere³⁵. Briefly, PDMS (Dow Corning Sylgard 184 Silicone) pre-polymer is poured onto the surface of a Fresnel lens and a diffraction grating separately. After baking the PDMS for 12 hours at 60 °C, PDMS is cured completely and a negative Fresnel lens mold and a negative grating mold are made. Afterward, a G-Fresnel can be fabricated by sandwich the PDMS pre-polymer between the two negative molds followed by curing. Raw Bayer pattern images captured by the CMOS detector are first converted to grey level two-dimensional spectral image by summing up the red, green and blue pixel values. The grey level spectral images are then summed by column (along the slit direction) to obtain the final one-dimensional spectral data. The spectrometer enclosure case is fabricated by three-dimensional printing, which has pre-designed slots for positioning and aligning all the components. Slot positions, such as these for holding the input slit, the G-Fresnel, and the camera detection system are first



Figure 2. Schematic diagram of the diffuse reflectance spectroscopy experimental system by using the G-Fresnel smartphone spectrometer.

determined on an optical bench. Trial and error is then used to arrive at the final design parameters of the case. Once the case is fabricated, only minor alignment is required. Control electronics is implemented so that the spectrometer can connect to a smartphone using the microUSB charging port, common across Android smartphone platforms. Power is also provided to the CMOS camera from the microUSB connection. An Android APP is developed to configure and control the spectrometer, as well as to transfer the spectral data to the phone for analysis. The Android APP was developed in Eclipse IDE for Java developers. It allows a user to set the integration time, initiate a measurement, store the measured data, display the spectra, and perform simple analysis on the results. Specifically, a spectral capture is initiated by pressing a button on the smartphone screen, which sends a signal to the microUSB back to the APP. The APP stores the data, processes it, and plots the spectrum on the screen. A photo of the smartphone spectrometer is presented in Fig. 1b, which shows the device attached to a smartphone and the APP measurement interface. Wavelength calibration is performed by identifying the wavelength peaks of a calibration lamp and the corresponding pixel positions of the spectral lines captured. The relationship is fitted with a linear curve and the parameters are stored on the phone.

Diffuse reflectance spectroscopy system. The DRS experimental setup is illustrated in Fig. 2. Light from a broadband tungsten halogen lamp (HL-2000-HP, Ocean Optics) is coupled into an Ocean Optics fibre probe, which consists of 6 multimode fibres surrounding a single fibre in the center (see inset in Fig. 2). The fibre core size is 400 μ m, and the diameter of each fibre and the seperation between the centers of neighboring fibres are ~480 μ m. The six peripheral fibres are used for illumination while the central fibre collects the diffusely reflected light and delivers it to the G-Fresnel smartphone spectrometer.

Tissue phantom preparation. The liquid tissue phantom samples are mixtures of human hemoglobin (H0267, Sigma-Aldrich Co. LLC) as the absorber and 1- μ m polystyrene microspheres (07310–15, Polysciences, Inc.) as the scatterers dispersed in water. The polystyrene microspheres are used to simulate tissue scattering. Hemoglobin, as the sole absorber in this sample, determines the absorption of the sample following Beer-Lambert Law. Here, we prepared 16 phantoms, which cover a hemoglobin (100% oxidized) concentration range from 5.39 to 36.16 μ M in water, by fixing the number of microspheres and titrating the hemoglobin.

Tissue phantom measurement. In order to measure the diffuse reflectance spectrum, the tip of the fibre probe was brought into contact with the phantom surface and a magnetic stirrer was used to ensure a uniform colloidal suspension of the microspheres throughout the measurements. Individual spectrum was measured from each phantom using the fibre probe and the G-Fresnel smartphone spectrometer with an integration time of 3.6 seconds. Immediately after all phantom measurements, the probe tip was placed in contact with a reflectance standard (Spectralon[®] SRS-99, LabSphere, Inc.) that has a flat reflectivity across all wavelengths. Experiments were performed at room temperature with ambient light on. Initially, a background measurement was also performed, which accounts the contribution from ambient light. This background spectrum was then subtracted

Phantom	Hb concentration (µM)	Expected μ_a (cm ⁻¹)	Expected µ's (cm ⁻¹)
1	5.39	0.46	12.55
2	8.80	0.75	11.95
3	11.90	1.02	11.41
4	14.74	1.26	10.92
5	17.34	1.48	10.46
6	19.73	1.69	10.05
7	21.94	1.87	9.66
8	23.98	2.05	9.31
9	25.89	2.21	8.98
10	27.66	2.36	8.67
11	29.31	2.50	8.38
12	30.86	2.64	8.11
13	32.31	2.76	7.86
14	33.67	2.88	7.62
15	34.95	2.99	7.40
16	36.16	3.09	7.19

Table 1. Summary of the expected hemoglobin concentrations as well as the mean absorption and scattering coefficients (averaged over 430–630 nm) of the 16 phantoms.

from the measured diffuse reflectance and reference spectra. For comparison, a benchtop DRS system¹⁷ was also utilized to obtain a diffuse reflectance spectrum from each of the same 16 phantoms.

Retrieval model. A Monte Carlo (MC) inverse model³⁷ of reflectance is used to extract the phantom absorption coefficient $\mu_a(\lambda)$ and reduced scattering coefficient $\mu_s'(\lambda)$ as well as hemoglobin concentrations (100% HbO₂) from the phantom spectrum between 430–630 nm. The wavelength range of 430–630 nm was selected to include the hemoglobin absorption peaks while avoiding those with low signal-to-noise ratio (e.g., between 400–430 nm). In each MC inversion, one of the phantoms was selected as the reference (with known $\mu_a(\lambda)$, $\mu_s'(\lambda)$ and reflectance spectrum) to analyze all the other phantoms. This process was repeated 16 times, in which each of the phantoms was used as the reference once. Finally the mean $< \mu_a(\lambda) >$ and $< \mu_s'(\lambda) >$ and their standard deviations (or error bars) across the 16 reference phantoms were calculated.

Results

Wavelength resolution characterization of the G-Fresnel smartphone spectrometer. The spectral resolution of the G-Fresnel smartphone spectrometer is characterized through a tunable narrow-band (<1 nm) calibration source created by passing a supercontinuum beam through a monochromator (PI Acton SpectraPro). The wavelength is increased at a step of 50 nm. At each wavelength, a spectrum of the calibration source is captured by using the smartphone spectrometer. Results are presented in Fig. 1c. Full-width at half-maximum (FWHM) values are determined at several representative wavelengths and are listed in the figure, indicating a wavelength resolution of about 5 nm. The spectrometer can be used within a wavelength range covering the visible and the near infrared region (~400 nm-1000 nm, limited by the sensitivity of the CMOS detector).

Tissue phantom experiment. To examin the feasibility of quantitative measurement of hemoglobin concenetration and scattering in tissue, we have used the DRS experimental system to perform proof-of-concept studies on the liquid tissue phantoms.

Two key parameters for DRS, absorption and reduced scattering coefficient, are pre-calculated. The absorption coefficient ($\mu_a(\lambda)$) was determined from a spectrophotometer (Lambda 35, PerkinElmer Inc.) measurement of a diluted hemoglobin stock solution and the reduced scattering coefficients ($\mu_s'(\lambda)$) was calculated using the Mie's theory³⁸ for known size, density and refractive index of the scatters. The phantom hemoglobin concentrations and the expected μ_a and μ_s' averaged over the wavelength range of 430–630 nm are presented in Table 1.

Figure 3 shows the diffuse reflectance spectra measured from all phantoms, obtained by dividing the background subtracted diffuse reflectance spectrum with the background subtracted reference spectrum. The three major absorption bands (Sorét, α and β bands) of oxyhemoglobin are clearly observed. The measured reflectance at the absorption peaks decreases with increasing hemoglobin concentrations as expected.

By using the Monte Carlo inversion model, both the absorption and the reduced scattering coefficients were obtained from the diffuse reflectance spectra data measured with the smartphone spectrometer and the benchtop spectrometer. The results are plotted against expected values in Fig. 4a and b. The percent errors, which are the relative difference between the extracted and expected values in μ_a and μ_s' , are also computed and listed in Fig. 4. We have obtained retrieval results with errors of 9.2% and 8.1% for absorption and reduced scattering coefficients, respectively, using the MC retrieval method. From the extracted absorption coefficients, the corresponding hemoglobin concentrations are calculated and presented in Fig. 4c. Typical errors for benchtop visible DRS systems that are used for preclinical and clinical studies are $5-10\%^{21,39,40}$. The error in absorption for this smartphone spectrometer is on the higher end; but it is still within an acceptable range. The extracted μ_a and μ_s' showed a



Figure 3. Measured diffuse reflectance spectra at different hemoglobin concentration.



Figure 4. Comparison of retrieved hemoglobin concentration, absorption coefficient and scattering coefficient using the G-Fresnel smartphone spectrometer and these using a benchtop DRS system.

general trend of higher deviation from the expected values at the two ends, which was due to fewer neighboring phantoms with close μ_a and μ_s ' values used for reference than those in the middle.

Our smartphone spectrometer based measurements compare favorably with the benchtop DRS system developed using a traditional spectrometer (AvaSpec-ULS2048-USB2-RM)²¹, which has errors of 6.1% and 2.2% for retrieved absorption and reduced scattering coefficients, respectively. Note that for tissue studies, only the reference phantom that produces the least error is selected as a reference to extract tissue hemoglobin concentrations.

Discussion

Smartphone based spectrometers have in recent years attracted significant interests thanks to the ubiquitousness of smartphones and the proven capability of optical spectroscopy for molecular sensing. The use of the G-Fresnel device, with dual-functionalities of focusing/collimation and wavelength dispersion, can significantly simplify the design, reduce the cost, weight and size, while maintains the high performance of optical spectrometers. Our development of an "add-on" spectrometer system attachable to a smartphone leverages the same low-cost and high-performance mobile CMOS camera technology, while allowing for direct access to raw spectral data as opposed to processed ones from a built-in image sensor on a smartphone. In addition, the broadband light source can be replaced with a cost efficient micro-lamp or light emitting diode and included with the spectrometer to make a complete spectrophotometer system, resulting in a compact DRS system suitable for POC applications. The DRS retrieval errors may be further decreased by improving the wavelength resolution of the smartphone spectrometer, which can be achieved by reducing the slit size, utilizing a G-Fresnel with finer grating period, and optimizing the optical alignment. Theoretical analysis previously indicates that 1 nm resolution can be achieved with a miniature G-Fresnel³³.

We note that the use of tissue phantoms to estimate the performance of the smartphone spectrometer is subject to several limitations. Firstly, the liquid phantoms are treated as a homogeneous medium, which may not be an ideal model for heterogeneous or layered biological tissues. The DRS only measures volume-averaged absorber concentrations and scattering properties in tissues. It is challenging to quantify non-uniformities at high resolution. Secondly, the blood of *in vivo* tissues generally includes both oxy- and deoxy-hemoglobin molecules; but

the hemoglobin molecules in the liquid phantoms used in the current experiments are mostly saturated with oxygen, which cannot be used to determine the accuracy for tissue oxygenation measurement. Finally, in the visible to near infrared wavelength range major scatterers are cell nuclei, mitochondria and collagen fibres with a size ranging from 10 micrometers to sub-micrometers. Therefore, future *in vivo* studies are necessary in order to fully evaluate the performance of the smartphone spectrometer for quantifying tissue hemoglobin contents.

Conclusion

In this work, we demonstrated a G-Fresnel spectrometer that can be attached to a smartphone, and based on it a DRS system for quantitative diffuse reflectance measurement. The G-Fresnel spectrometer offers a compact package yet maintains ~ 5 nm wavelength resolution across the visible and the near-infrared region. Further, we have conducted measurement of the hemoglobin concentrations and reduced scattering coefficients of tissue phantoms using the smartphone spectrometer based DRS system. In comparison to a benchtop spectrometer based DRS system, our system demonstrates the feasibility to perform quantitative hemoglobin measurements with a comparable retrieval error of 9.2%. By virtue of the compact size, portability, and low cost of the demonstrated system, the efficacy of optical spectroscopy for quantitative measurement, and the ease of data collection, management, and computing afforded by smartphone technology, new POC opportunities with broad accessibility could thus be enabled.

References

- 1. Johnson, N. J. et al. The association between hemoglobin concentration and neurologic outcome after cardiac arrest. J. Crit. Care 36, 218–222 (2016).
- 2. Kwong, T. C. et al. Differentiation of tumor vasculature heterogeneity levels in small animals based on total hemoglobin concentration using magnetic resonance-guided diffuse optical tomography in vivo. Appl. Opt. 55, 5479–5487 (2016).
- Panwar, B. et al. Hemoglobin Concentration and Risk of Incident Stroke in Community-Living Adults. Stroke 47, 2017–2024 (2016).
 Suemori, T. et al. Cerebral oxygen saturation and tissue hemoglobin concentration as predictive markers of early postoperative
- outcomes after pediatric cardiac surgery. *Pediatr. Anesth.* **26**, 182–189 (2016). 5. Sukrat, B. *et al.* Hemoglobin concentration and pregnancy outcomes: a systematic review and meta-analysis. *Biomed Res. Int.* **2013**,
- (2013). 6. Leal-Noval, S. R., Munoz-Gomez, M. & Murillo-Cabezas, F. Optimal hemoglobin concentration in patients with subarachnoid
- hemorrhage, acute ischemic stroke and traumatic brain injury. *Curr. Opin. Crit. Care* 14, 156–162 (2008). 7. Hoff, C. M. Importance of hemoglobin concentration and its modification for the outcome of head and neck cancer patients treated
- with radiotherapy. *Acta Oncol. (Madr).* **51**, 419–432 (2012). 8. Dellas, A. *et al.* Angiogenesis in cervical neoplasia: microvessel quantitation in precancerous lesions and invasive carcinomas with
- clinicopathological correlations. *Gynecol. Oncol.* 67, 27–33 (1997).
 P. Flexman, M. L. *et al.* Optical biomarkers for breast cancer derived from dynamic diffuse optical tomography. *J. Biomed. Opt.* 18, 96012 (2013).
- 10. Pogue, B. W. et al. Characterization of hemoglobin, water, and NIR scattering in breast tissue: analysis of intersubject variability and menstrual cycle changes. J. Biomed. Opt. 9, 541–552 (2004).
- 11. Tromberg, B. J. et al. Non-invasive in vivo characterization of breast tumors using photon migration spectroscopy. Neoplasia 2, 26–40 (2000).
- 12. Müller, M. G. *et al.* Spectroscopic detection and evaluation of morphologic and biochemical changes in early human oral carcinoma. *Cancer* **97**, 1681–1692 (2003).
- Amelink, A. *et al.* Non-invasive measurement of the morphology and physiology of oral mucosa by use of optical spectroscopy. *Oral Oncol.* 44, 65–71 (2008).
- 14. Sharwani, A. et al. Assessment of oral premalignancy using elastic scattering spectroscopy. Oral Oncol. 42, 343–349 (2006).
- 15. Subhash, N. *et al.* Oral cancer detection using diffuse reflectance spectral ratio R540/R575 of oxygenated hemoglobin bands. *J. Biomed. Opt.* **11**, 14018 (2006).
- Chang, S. K. et al. Combined reflectance and fluorescence spectroscopy for in vivo detection of cervical pre-cancer. J. Biomed. Opt. 10, 24031–2403111 (2005).
- Chang, V. T.-C., Bean, S. M., Cartwright, P. S. & Ramanujam, N. Visible light optical spectroscopy is sensitive to neovascularization in the dysplastic cervix. J. Biomed. Opt. 15, 57006 (2010).
- 18. Mirabal, Y. N. et al. Reflectance spectroscopy for in vivo detection of cervical precancer. J. Biomed. Opt. 7, 587-594 (2002).
- Palmer, G. M. et al. Quantitative diffuse reflectance and fluorescence spectroscopy: tool to monitor tumor physiology in vivo. J. Biomed. Opt. 14, 24010 (2009).
- Phelps, J. E., Vishwanath, K., Chang, V. T. C. & Ramanujam, N. Rapid ratiometric determination of hemoglobin concentration using UV-VIS diffuse reflectance at isosbestic wavelengths. Opt. Express 18, 18779–18792 (2010).
- 21. Yu, B., Shah, A., Nagarajan, V. K. & Ferris, D. G. Diffuse reflectance spectroscopy of epithelial tissue with a smart fiber-optic probe. *Biomed. Opt. Express* 5, 675–689 (2014).
- Stratonnikov, A. A. & Loschenov, V. B. Evaluation of blood oxygen saturation *in vivo* from diffuse reflectance spectra. J. Biomed. Opt. 6, 457–467 (2001).
- Wang, H.-W. et al. Diffuse reflectance spectroscopy detects increased hemoglobin concentration and decreased oxygenation during colon carcinogenesis from normal to malignant tumors. Opt. Express 17, 2805–2817 (2009).
- 24. Smith, Z. J. *et al.* Cell-phone-based platform for biomedical device development and education applications. *PLoS One* **6**, e17150 (2011).
- 25. Gallegos, D. et al. Label-free biodetection using a smartphone. Lab Chip 13, 2124-2132 (2013).
- 26. Wang, S. X. & Zhou, X. J. Spectroscopic sensor on mobile phone. (2008).
- Long, K. D., Yu, H. & Cunningham, B. T. Smartphone instrument for portable enzyme-linked immunosorbent assays. *Biomed. Opt. Express* 5, 3792–3806 (2014).
- 28. Arafat Hossain, M. et al. Combined 'dual' absorption and fluorescence smartphone spectrometers. Opt. Lett. 40, 1737–1740 (2015).
- 29. Zhang, C., Kim, J. P., Creer, M., Yang, J. & Liu, Z. A smartphone-based chloridometer for point-of-care diagnostics of cystic fibrosis. *Biosens. Bioelectron.* (2017).
- Das, A. J., Wahi, A., Kothari, I. & Raskar, R. Ultra-portable, wireless smartphone spectrometer for rapid, non-destructive testing of fruit ripeness. Sci. Rep. 6, 32504 (2016).
- 31. Hossain, M. A., Canning, J., Cook, K. & Jamalipour, A. Optical fiber smartphone spectrometer. Opt. Lett. 41, 2237–2240 (2016).
- 32. Yang, C., Edwards, P. S., Shi, K. & Liu, Z. A compact optical spectrometer based on a single-grating Fresnel diffractive optical element. in *SPIE Defense, Security, and Sensing* 83740Y–83740Y (International Society for Optics and Photonics (2012).

- Yang, C., Edwards, P., Shi, K. & Liu, Z. Proposal and demonstration of a spectrometer using a diffractive optical element with dual dispersion and focusing functionality. Opt. Lett. 36, 2023–2025 (2011).
- Yang, C., Edwards, P., Shi, K. & Liu, Z. Hybrid Diffractive Optical Element Based Spectrometer. in CLEO: Science and Innovations JWA64 (Optical Society of America, 2011).
- 35. Yang, C., Shi, K., Edwards, P. & Liu, Z. Demonstration of a PDMS based hybrid grating and Fresnel lens (G-Fresnel) device. *Opt. Express* 18, 23529–23534 (2010).
- 36. Zhang, C. et al. G-Fresnel smartphone spectrometer. Lab Chip 16, 246-250 (2016).
- Palmer, G. M. & Ramanujam, N. Monte Carlo-based inverse model for calculating tissue optical properties. Part I: Theory and validation on synthetic phantoms. Appl. Opt. 45, 1062–1071 (2006).
- 38. Bohren, C. F. & Huffman, D. R. Absorption and scattering of light by small particles. (John Wiley & Sons, 2008).
- Bender, J. E. et al. A robust Monte Carlo model for the extraction of biological absorption and scattering in vivo. IEEE Trans. Biomed. Eng. 56, 960–968 (2009).
- Alerstam, E., Andersson-Engels, S. & Svensson, T. Improved accuracy in time-resolved diffuse reflectance spectroscopy. Opt. Express 16, 10440–10448 (2008).

Acknowledgements

ZL acknowledges the support of the Vodafone Americas Foundation and the Penn State Humanitarian Materials project.

Author Contributions

Z. Liu, P. Edwards and B. Yu conceived the project. P. Edwards, C. Zhang, B. Zhang built and characterized the device. P. Edwards, X. Hong and V.K. Nagarajan conducted the tissue phantom experiments. X. Hong, V.K. Nagarajan and B. Yu performed the retrieval of tissue phantom parameters from the measured diffuse reflectance spectra. All authors contributed to the preparation of the paper.

Additional Information

Competing Interests: ZL and PE have a financial interest in Atoptix Inc, a company which could potentially benefit from the results of this research. This interest for ZL has been reviewed by the University in accordance with its Individual Conflict of Interest policy, for the purpose of maintaining the objectivity and the integrity of research at The Pennsylvania State University.

Publisher's note: Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit http://creativecommons.org/licenses/by/4.0/.

© The Author(s) 2017